

Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan
Grievance- M/s Medequips Vs M/s Eastern Medical Technology Services

ADP Scheme "Provision of Radiological Equipment in 120-Bedded Emergency Block at Sheikh Zayed Medical Complex, Rahim Yar Khan"

Sr. No.	Observations	Date	Status of Observation	Encloser	Remarks of the Committee
A	M/s Eastern Medical Technologies (EMTS) has been commercially disqualified in Punjab Health Deptt. Projects on the basis of dissolution of their partnership deed. Copies of few evaluation findings by the Govt. of the Punjab are attached for quick reference		Detail as given below:-	Annex-A	
	i- Project of Tertiary care Hospitals by SHC & ME Department, GoP.	21/2/2018	Enclosed Minutes of Meeting are not Complete (decision is not available)		<i>Objection cannot be verified as decision page is not included in docu</i>
	ii- Project of IDAP (Infrastructure Development)	2/4/2018	In 2016 IDAP announced that EMTS has no leagal status as a partnership firm from 7th May 2016 & asked competent authority for black listing on Misinterpretation. (Minutes of meeting are not complete).		<i>objection accepted</i>
	iii- Project of Strategic Management Unit (SMU)	9/6/2018	Purchase of Patient Ventilators: EMTS stood non-compliant due to dissolution of partnership		<i>objection accepted</i>
	iv- Final Notification Issued by SHC & ME Deptt. GoP	5/4/2019	Purchase of Patient Monitors: EMTS stands non-responsive and any action against EMTS will be on the decision of Hon'able Lahore high court under writ petition NO. 2595/18		<i>objection accepted</i>

JAM TAHIR RASOOL
 Bio Medical Engineer
 Sheikh Zayed Medical College
 Hospital Rahim Yar Khan

Dr. Zaheer Mustafa
 MBBS, FCPS
 Associate Professor Radiology
 Sh. Zayed Medical College
 Rahim Yar Khan

[Handwritten signatures and initials: JAM TAHIR RASOOL, Dr. Zaheer Mustafa, and others]

Stamp: **RAHIM YAR KHAN**
 Sheikh Zayed Medical College
 Hospital Rahim Yar Khan

Sr. No.	Observations	Date	Status of Observation	Encloser	Remarks of the Committee
B	As per bidding documents clause # 15.3 (d); the bidder should have three year Experience in the Market. Please check Form-C for the confirmation of their partnership deed with registerar office	1996-97	M/s EMTS as per Taxpayer Registration certificate have been in the marketplace since 1996-97. However, <u>its partnership ends on 7th May 2016. Copy of Form-C is enclosed.</u>	Annex-B	<i>Objection accepted</i>
C	Please refer bidding document clause # 15.3 (j); which bounds the firms to provide list of installation dates with contact number received from private & public hospitals over last two years.		Copies of said supply orders & details are enclosed provided in the bid of M/s EMTS	Annex-C	<i>No orders in 2017 & 2018. Objection accepted</i>
D	M/s EMTS don't have a professional Service support for such a complex installation and after sale service; Since it's a purchase of highly sophisticated equipment; qulaified trained engineers, sufficient stock inventory of spare parts and specilized workshop facilities equipped with advance test tools etc. <u>are mandatory requirement that can be easily examined by arranging physical Inspection.</u>		List of Technical Staff & Tools is enclosed, provided in the bid of M/s EMTS	Annex-D	<i>objection Rejected as cost is provided by EMTS.</i>
E	M/s HITACHI has a very bad reputation in punjab Govt. Projects. They installed only one 128 Slice CT Scanner in Govt. Mayo Hospital, which was not properly commissioned and has a lot of troubles since installion. The latest status of this machine may kindly be checked from the		copy of IDAP decision is enclosed submitted by M/s Medequips		<i>Objection accepted poor performance of installed machine in Mayo in not properly working & EMTS</i>
F	M/s HITACHI has many non-Exclusive Distriubtors including M/s Ontech Medical, M/s Biotech Pakistan, M/s SK ent. M/s Ittefaq Sons & M/s Raditech, therefore, reliability of professional back support is not sure		M/s Eastern Medical Technology Services have authorization for CT Scan, MRI & Ultrasound (copy enclosed provided in the bid of M/s EMTS	Annex-E	<i>objection accepted</i>

JAMIL TAHIR RASOOL
 Medical Engineer
 Sh. Yayed Medical College
 Hospital Rawalpindi
 Hospital Lahore

Dr. Zahoor Mushtaq
 MBBS, FCPS
 Associate Professor Radiology
 Sh. Yayed Medical College
 Hospital Lahore

services there. It is telephonically verified from Dr. Khurshid, Associate Professor Radiology, Hospital Lahore by Prof. Dr. Saeed Akhtar Malik

Khawaja
 Dr. Khurshid
 Dr. Saeed Akhtar Malik
 Dr. Zahoor Mushtaq

Sr. No.	Observations	Date	Status of Observation	Encloser	Remarks of the Committee
	The Qouted Model by M/s EMTS "Scenaria View" has following Major Technical deviations:		as mentioned below:-		
G	Tender asks for "workstation will be supplied from manufacturer i- of CT to ensure similar work-flow"; whereas, HITACHI offers third party workstation which is a serious deviation.		Yes HITACHI have a partnership with TeraRecon, Inc for workstation		Objection Accepted
G	HITACHI's Model Scenaria View doesn't comply with Int. standard FDA (510K), which is a mandatory requirement of such high tech equipment, where there is a great concern of radiation ii- emission. Please note that Govt. of the Punjab Health Deptt. Always prefer the radiation equipment must have FD 510K certificate because FDA has very strict rules in observing radiation factors		Its not mandatory to have FDA as per standard bidding documents of Health Department, rather any two from FDA, EC, MDD & MHLW. Copy of standard bidding document enclosed.		Objection Rejected as EMTS have certificates from CE EC & MHLW. <i>was</i>
G	Tender asks for "Dual energy scanning system with blending software" whereas, Scenaria View uses and older scan/ scan dual energy technology to acquire the whole data in two gantry rotation, while Medequips offers advance KV switching iii- Technology to acquire the whole data in single gantry rotation. Please note that there are many disadvantages of older scan technology including high radiation dosage, more uses of contrast media, moving artifacts, high chances of repeat scan etc.		Dual Energy is available, however, single or double rotation is not verified		Objection Rejected

Decision:

Technically rejected on the basis of observation. A (i, ii, iii) B, C, E, F & G (i) was

JAM TAHIR RASOOL
Bio Medical Engineer
Shahkhan Zayed Medical College
Hospital Rahim Yar Khan

Dr. Zaheer Mustafa
MBBS, FCPS
Associate Professor Radiology
Sh. Zayed Medical College
Rahim Yar Khan

PRINCIPAL
Shahkhan Zayed Medical College
Rahim Yar Khan

was

Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan

Grievance- M/s Shirazi Trading

ADP Scheme "Provision of Radiological Equipment in 120-Bedded Emergency Block at Sheikh Zayed Medical Complex, Rahim Yar Khan"

Sr. No.	Observations	Date	Clarification By Shirazi Trading	Remarks of the Committee
A	Country of origin is China		<p>As per Clause 4.L of Standard Bidding Document issued by Health Department & your hospital explaining Country of origin'</p> <p><u>"Country of origin should be of USA, Europe & Japan, unless otherwise any other country of manufacturer is mentioned in the specifications. However, country of origin of equipment could be from any geographical region of the world as per laws of Pakistan."</u> Document issued by Health Department & your hospital explaining Country of origin'.</p> <p>M/s Shirazi comply the tender specs and were declared ResPonsive in the Rawalpindi Institute of Cardiology & D.G Khan Medical College Dera Ghazi Kha n Tenders.</p>	/
B	1ms shortest exposure time		<p>0.0002 sec shortest exposure time sufficient to manage any motion artifact.</p> <p>- Current , Voltage and exposure time rating of X-raY tube plays vital role in dose and quality imaging.</p> <p>-Attached research PaPers and ACR guidelines clearly shows recommended exposure time , voltage and current values for most radiograPHY examinations, all of them are within XR 646 caPabilities.</p> <p>-We would request to Please accept this as a mlnor deviation because as Per recommendation it will not affect any major requirement.</p>	<p><i>Objection rejected as exposure time is out of specification Original decision is maintained.</i></p>

Abdul Rauf

JAMTAHIR RASOOL
 B.E. Medical Engineer
 Sheikh Zayed Medical College
 Hospital Rahim Yar Khan

Dr. Zaheer Mustafa
 Associate Professor Radiology
 Sheikh Zayed Medical College
 Hospital Rahim Yar Khan

Dr. Zaheer Mustafa
 MBBS, FCPS
 Associate Professor Radiology
 Sheikh Zayed Medical College
 Hospital Rahim Yar Khan


Abdul Rauf

C	Matrix size aPProximately 2800 x 2800 or better	<p>Contradiction in tender specification , a detector with 200 micro size(as mentioned in tender) can have maximum resolution of 2022 X 2022. Further, GE Healthcare have dedicated wireless FLASHPAD Detector having2022X2022 with highest DQE 68%.</p> <p>-As per tender specs, requirement is for 200micron pixel size for la rge detector but 2800X2800 resolution is not possible with req uested pixel size.</p> <p>- This means fulfilling one parameter of tender specs will result in deviation in values of second Parameter.</p>	<p>Objective accepted Out of specification & So original design of Rejectin Stand Valid.</p>
---	---	---	---


Decision:

Stand Rejected.

Technically M/s Shirazi Trading Stand rejected


JAM TAHIR RASOOL
Bio Medical Engineer
Sheikh Zayed Medical College
Hospital Rahim Yar Khan


Dr. Zaheer Mustafa
MBBS, FCPS
Associate Professor Radiology
Sh. Zayed Medical College


Sh. Zayed Medical College Hospital
Rahim Yar Khan

