

SHEIKH ZAYED HOSPITAL, RHAIM YAR KHAN

BIDDER WISE TECHNICAL COMPARATIVE STATEMENT FOR THE BULK PURCHASE OF BLOOD GAS ANALYZER ITEMS FY 2021-22

Sr#	Detail	M/S Hoora Pharma	M/S BioMadics	M/S A.M Healthcare	M/S MediTec
		1	2	3	4
1	Original receipt for purchase of tender.	Attached	Attached	Attached	Attached
2	Attested copy of CNIC	Attached	Attached	Attached	Attached
3	Acceptance of terms and condition, tender documents duly signed and stamped.	Signed & Stamp	Signed & Stamp	Signed & Stamp	Signed & Stamp
4	Valid Manufacturer's Authorization	Attached	Attached	Attached	Attached
5	An affidavit on stamp paper of Rs.100/- submitting following clauses: i) replacement of unconsumed / expired / substandard spurious drugs / stocks free of cost, ii) that the firm is never blacklisted on any grounds whatsoever. iii) Price Reasonable certificate. (iv) Certificate that prices are not more than trade price.	Attached	Attached (Rs. 50/- instead of Rs. 100/-)	Attached	Attached
6	An affidavit on stamp paper submitting that the price quoted to this institute against the quoted items mentioned in the bid are not more than the prices charged from any Purchase Organization in the country and in case of discrepancy the bidder hereby undertakes to refund the price charged in excess".	Attached	Attached (Rs. 50/- instead of Rs. 100/-)	Attached	Attached
7	Affidavit on Original Letter Head (as per specimen proforma attached on Page No.27)	Attached	Attached	Attached	Attached
8	Call deposit / Bid Security required (2% of estimated price)	Attached	Attached	Attached	Attached (Not Visible)
9	National Tax Number (NTN)	Attached	Attached	Attached	Attached
10	General Sale Tax Number Certificate (STN)	Attached	Attached	Attached	Attached
11	Professional Tax	Attached	Attached	Attached	Attached
12	Bank Statement last one year.	Attached	Attached	Attached	Attached
13	Experience Certificate / Past Performance Certificate	Attached (07 Years)	Attached (06 Years)	Attached (08 Years)	Attached (07 Years)
14	Sample provided as per Proforma attached (Page No.29)	Mentioned against each item in Item wise Comparative Statement	Mentioned against each item in Item wise Comparative Statement	Mentioned against each item in Item wise Comparative Statement	Mentioned against each item in Item wise Comparative Statement

Sr#	Detail	M/S Hoora Pharma	M/S BioMadics	M/S A.M Healthcare	M/S MediTec
		1	2	3	4
15	The Original price List of the firm indicating the inclusion of the item / product of quoted item (Latest)	Attached	Attached	Not Attached	Not Attached
16	The bidder is required to provide Financial Proposal with the name of items, tender number and serial number in the exact manner as quoted in the Technical Proposal	Followed	Followed	Followed	Followed
17	Price should not be mentioned on technical bid.	Followed	Followed	Followed	Followed
18	Quality Certificates (ISO/CE/IVD/ FDA/ WHO Recommendation, any other)	Mentioned against each item in Item wise Comparative Statement	Mentioned against each item in Item wise Comparative Statement	Mentioned against each item in Item wise Comparative Statement	Mentioned against each item in Item wise Comparative Statement
19	Valid Import License (<i>issued by DRAP in case of Importers</i>)	Attached	Attached	Attached (Importer's License)	Attached
20	Sole Agency Certificate / Agreement with foreign Principal (in case of Importer) Translated in English.	Attached	Attached	Attached (Authority Letter)	Attached
21	Letter of Intention (as per specimen proforma attached)	Attached	Attached	Attached	Attached
22	Manufacturer, Model & Country of Origin of Quoted Instrument in case if any.	Mentioned against each item in Item wise Comparative Statement	Mentioned against each item in Item wise Comparative Statement	Mentioned against each item in Item wise Comparative Statement	Mentioned against each item in Item wise Comparative Statement
23	Remarks	Responsive	Non responsive Subject to provision of following documents: 1) Affidavit at Sr. 5&6 attached of Rs. 50/- Instead of Rs. 100/-	Responsive	Non responsive Subject to provision of Visible copy of CDR

MEDICAL SUPERINTENDENT
Sheikh Zayed Hospital,
Rahim Yar Khan

SHEIKH ZAYED MEDICAL COLLEGE/ HOSPITAL RAHIM YAR KHAN

Item-Wise Technical Comparative Statement for the Bulk Purchase of Blood Gas Analyzer (ABGs) for Financial Year 2021-22.

BID EVALUATION CRITERIA

Tender Sr. No. 1-5

Three Automated Blood Gass Analyzer with UPS (2 KVA PURE SINE WAVE) on reagent rental basis:

Throughput minimum 30 test/hour, minimum TAT 50 second. Principle methodology: Potentiometric, Enzyme/Amperometric, indirect ISE. Sample type: Heparinized, Whole Blood, Arterial, Venuous, Mixed Venuous (Syring and Capillary Mode).

Name of Item: Measured Parameters: pH,PCO2 ,PO2, SO 2%, Hb,HCT,Na,K,Cl,Ca , Calculated Parameter: Base excess of blood, Bicarbonate HCO3, pH,PCO2 ,PO2, (Corrected to patient temprature). Measuring range: PCO2 200-300mmHg, PO2 0-800mmHg, pH 6.5-8.00. Automatic 2-point Calibration with Manual Calibration, Built-in Display, Printer. Complete with All Standared and Operational Accessories, Reagents, Calibrators, Electrodes, Cleaner, Printer Roll and Controls for Startup Kit. CE/FBA/ISO Approved (Make: USA, Europe, Japan, UK) with or without POCT Devices Two in No. at least 15000 Tests

Sr.#	Name of Firm	Brand / Manufacturer Name	Quoted Pack Size	Ordinary Criteria of Firm/ Bidder													Compulsory Criteria of Product					Approval of samples by TSC/ End User & Previous Experience, (Approved "A" Rejected "R",)	Obt. Marks	Technical Eligibility of Product (Responsive / Non-Responsive)	Remarks
				a) (Past Performance of the Bidder "PP" (Last two years) b) Market Experience of quoted Product "MPE" c) Credibility & Certification of Manufacturer "CrC" (ISO, CE, IVD, FDA etc), d) Financial Status of Bidder "FS" (Audit Report or Balance Sheet) e) Company Profile "CP" (Bio-medical engineer, software engineer, Application Engineer, Services & Backup, Supporting Staff List) f) Financial Capacity of the Bidder / Firm (Annual Turnover of last financial year)													a) Authority Letter "AL" of quoted brand b) Product Experience "PE" c) Free Sale Certificate "FSC" e) Sample Specification "SS"								
				(a) PP	(b) MPE	(c) CrC (10 marks)						(d) FS			(e) CP				(f) FC	AL Y/N	PE Y/N				
10	10	4	4	4	4	4	4	5	5	10	10	5	5	20											
1	M/S Hoora Pharma	Siemens Healthinner, UK	1) Buffer Solution (4 Bottles) ii) Wash Solution (4 Bottles) iii) Gas Cartridge Pack (1x1 No) 2) Cleaning Solution (FOC) 3) Quality Control (FOC) 4) All Consumables as required by the Instrument (FOC) 5) Thermal Printing Paper (3500 Tests)	8	10	4	0	4	4	0	5	5	10	10	5	5	20	Y	Y	Y	Y	Approved	90	Responsive	
2	M/S BioMadies	Gastat 720 Techno Medica, Japan	1) Cell Pack / Call Cartridge 2) Cleaning Solution 3) Quality Control (FOC) 4) All Consumables as required by the Instrument (FOC) 5) Thermal Printing Paper (3000 Tests within 30 days)	8	6	4	4	4	0	0	5	5	10	10	5	5	5	Y	Y	Y	Y	Approved	71	Responsive	

Sr.#	Name of Firm	Brand / Manufacturer Name	Quoted Pack Size	Ordinary Criteria of Firm/ Bidder													Compulsory Criteria of Product					Approval of samples by TSC/ End User & Previous Experience, (Approved "A" Rejected "R".)	Obt. Marks	Technical Eligibility of Product (Responsive / Non-Responsive)	Remarks			
				a) (Past Performance of the Bidder "PP" (Last two years) b) Market Experience of quoted Product "MPE" c) Credibility & Certification of Manufacturer "CrC" (ISO, CE, IVD, FDA etc), d) Financial Status of Bidder "FS" (Audit Report or Balance Sheet) e) Company Profile "CP" (Bio-medical engineer, software engineer, Application Engineer, Services & Backup, Supporting Staff List) f) Financial Capacity of the Bidder / Firm (Annual Turnover of last financial year)													a) Authority Letter "AL" of quoted brand b) Product Experience "PE" c) Free Sale Certificate "FSC" e) Sample Specification "SS"											
				(a) PP	(b) MPE	(c) CrC (10 marks)						(d) FS			(e) CP				(f) FC	AL Y/N	PE Y/N					FSC Y/N	SS Y/N	A/R
						ISO	CE	IVD	FDA	other	Audit Report	Bank Statement	Tax Return (3 Yrs)	Tech. Staff.	Service & Backup	Supporting Staff												
10	10	4	4	4	4	4	4	0	0	5	5	10	10	5	5	5												
3	M/S A.M Healthcare	NOVA, USA	1) Reagent Pack / State Profile i) Calibrator Catridge 2) Cleaning Solution 3) Quality Control (FOC) 4) All Consumables as required by the Instrument (FOC) 5) Thermal Printing Paper (1x400 Tests)	10	10	4	4	4	4	0	0	5	10	10	5	5	5	Y	Y	Y	Y	Rejected	76	Non-responsive	FSC Attached (Not Attested by Pak Embassy) (Rejected by the end user due to non-provision of Instrument)			
4	M/S Meditec Instruments Co.	Easystat Medica USA	i) Reagent Pack 1x850 ml 2) Cleaning Solution (90 ml) (1x6) 3) Quality Control (FOC) 4) All Consumables as required by the Instrument (FOC) 5) Thermal Printing Paper (No of tests not mentioned)	8	10	4	0	4	0	0	5	5	10	10	5	5	5	Y	Y	Y	Y	Rejected	71	Non-responsive	Rejected by the end user			

Total Marks
Qualifying Marks

100
60

MEDICAL SUPERINTENDENT
Sheikh Zayed Hospital,
Rahim Yar Khan