

Sheikh Zayed Medical College / Hospital, Rahim Yar Khan
Technical Evaluation Statement (Financial Year 2021-22) on 28-10-2021

ADP Scheme "Programme for Revamping of Emergencies of Tertiary Care Hospital in Punjab (Sheikh Zayed Hospital, Rahim Yar Khan)"

Sr.#	ITEM NO. 01E Transport Ventilator FOR PAEDTRIC	M/s Medical Equipment & System	M/s Hospicare Systems	M/s Eastern Medical	M/s Radiant Medical
	KNOCK DOWN CRITERIA (COMMERCIAL EVALUATION)				
1	Original Receipt for Purchase of Tender	Yes	Yes	Yes	Yes
2	An Affidavit (as per Specimen Performance on Stamp Paper of Rs. 100/-)	Yes	Yes	Yes	Yes
3	Copy of CDR/Bid Security	Yes	Yes	Yes	Yes
4	Bid Validity	Yes	Yes	Yes	Yes
5	Delivery Period	Yes	Yes	Yes	Yes
6	Valid Manufacturing Licence (Embassy Attested)	Yes (Online Verified Registration Certificate)	Yes (Online Verified Registration Certificate)	No Registration Certificate Provided and not a sole distributor	Yes (Online Verified Registration Certificate)
	Remarks (Responsive/Non-Responsive)	Responsive	Responsive	Non Responsive	Responsive
	Evaluation parameter (quoted times/Equipment)				
1	Exclusive /Sole authorization agreement from the Foreign Manufacturer.	Yes	Yes	Yes	Yes
2	Certificate from the Manufacturer that the after sales Services/backup Services shall be provided jointly with the local sole agent and in case of change of local agent, they will provide the after sale services themselves for through newly appointed agent for the period mentioned from the date of commissioning.	Yes	Yes	Yes	Yes
3	Submission of valid legally enforceable exclusive authorization letter of manufacturer assuring full guarantee and warranty (03 Years) obligations as per enclosed manufacturer authorized form with the bid document.	Yes	Yes	Yes	Yes
4	Valid Manufacturing License of quoted item Manufacturing embassy attested with certified documentary evidence of manufacturing site & location.	Yes	Yes	No Registration Certificate Provided and not a sole distributor	Yes
5	The medical equipment offered from foreign countries of USA, Europe and Japan shall be eligible to participate and must bear FDA510k, CE(MDD) or MHLW (Ministry of Health, Labor and Welfare) standard.	Yes	Yes	Yes	Yes
6	A Certificate from the manufacturer that the installation will be conducted in conformity with the system requirements by following the professional approach.	Yes	Yes	Yes	Yes
7	Satisfactory Past Performance of the bidder for quoted product. (02 years)	Yes	Yes	Yes	Yes
8	The firms shall also declare the make, model, country of origin and country of manufacturing of all standard accessories to be provided with the equipment.	Yes	Yes	Yes	Yes
9	The quoted model of imported product shall be available on the current official website of the manufacturer; otherwise the quoted product shall be considered obsolete/ redundant and will straight away be rejected. (Web link and Screenshot must be attached)	Yes	Yes	Yes	Yes
10	The bid must comply with the required technical specifications of the quoted single item/ complete package. Incomplete offer will straightaway be rejected.	Yes	Yes	Yes	Yes

Dr. Zenshan Khan
 M.B.B.S., F.C.P.S.
 Assistant Professor Anaesthesia ICU
 Sheikh Zayed Medical College/Hospital
 Rahim Yar Khan

ENG. AWAIS BIN TARIQ
 Bio Medical Engineer/Technical Officer
 Sh. Zayed Medical College/Hospital P.V.V.

DR. AKIF RAUF
 Deputy Medical Superintendent
 Sheikh Zayed Hospital Rahim Yar Khan

Prof. Dr. Muhammad Saleem
 MBBS, MRCPS, FRCPS (Gold Medalist)
 Head of Pediatric Department
 Sheikh Zayed Medical College/Hospital
 Rahim Yar Khan

Dr. M. M. C. Hospital, Rahim Yar Khan
 Head of Obs & Gynae Unit

Sr.#	ITEM NO. 01E Transport Ventilator FOR PAEDTRIC	M/s Medical Equipment & System	M/s Hospicare Systems	M/s Eastern Medical	M/s Radiant Medical
11	Acceptance of terms and condition of tender documents duly signed and stamped	Yes	Yes	Yes	Yes
	Remarks (Responsive/Non-Responsive)	Responsive	Responsive	Non Responsive	Responsive
	Evaluation Criteria (Score)	81	75	85	79
	TECHNICAL SPECIFICATIONS				
	Make	Zoll's	Medtronic Covidien AG group	SIARE ENGINEERING INTERNATIONAL GROUP	SIARE ENGINEERING INTERNATIONAL GROUP
	Model	EMV+731	Puritan Bennett™ 560	FALCO 202	ARIA 104
	Country of Manufacturer (Origin): USA, EU, Japan	USA	Ireland	Italy	Italy
	Evaluation Criteria Score				
	Microprocessor controlled ventilator with folloeing minimum modes.	Yes	Yes	Yes	Yes
	continuous Mandatory ventilator (CMV).	Yes	Yes	Yes	Yes
	synchronized Intermittent Mandatory ventilator (SIMV).	Yes	Yes	Yes	Yes
	Continuous positive airway pressure (CPAP)	Yes	Yes	Yes	Yes
	Pressure Support Ventilation	Yes	Yes	Yes	Yes
	Non Invasive Ventilation	Yes	Yes	Yes	Yes
	Tidal Volume 50-2000ml	Yes	Yes	Yes	Yes
	MONITORING DISPLAY:	-	-	-	-
	Graphic Screen 04" Screen or more	Yes	Yes	Yes	Yes
	Display parameters :- Airway pressure;Breath rate.	Yes	Yes	Yes	Yes
	Inspiratory time	Yes	Yes	Yes	Yes
	Exhaled minute volume I:E ratio	Yes	Yes	Yes	Yes
	Exhaled tidal Volume	Yes	Yes	Yes	Yes
	Trend Data	Yes	Yes	Yes	Yes
	Adjustable alarm for high/low respiratory rate:	-	-	-	-
	chargeable Built in battery backup,upto 4 hours and battery level indication.	Yes	Yes	Yes	Yes
	12V DC operations	Yes	Yes	Yes	Yes
	Complete with breathing circuits, hose, test lung & universal masks or other accessories required for proper functioning of the equipment.	Yes	Yes	Yes	Yes
	ACCESSORIES	-	-	-	-
	06 CIRCUITS (autoclave able)	Yes	Yes	Yes	Yes
	OPTIONAL				
	Remarks (Responsive/Non-Responsive)	Responsive	Responsive	Responsive	Responsive

Note:

M/S Eastren Medical is Non Responsive as per KNOCK DOWN CRITERIA.

Dr. Zeeshan Khan
M.B.B.S., F.C.P.S.
Assistant Professor Anaesthesia ICU
Sh. Zayed Medical College/Hospital
Rahim Yar Khan

Eng. AWAIS BIN TARIQ
Bio Medical Engineer/Technical Officer
Sh. Zayed Medical College/Hospital R.Y.K.

DR. AKIF RAU
Deputy Medical Superintendent
Sh. Zayed Medical College/Hospital R.Y.K.

DR. AKIF RAU
Deputy Medical Superintendent
Sh. Zayed Medical College/Hospital R.Y.K.

Prof. Dr. Muhammad Saleem
M.D., F.C.P.S., F.C.C.
Head of Pediatrics Department
Sh. Zayed Medical College/Hospital
Rahim Yar Khan

Prof. Dr. Nuzhat Rashid
Head of Obs & Gynae Unit-II
S.Z.M.C Hospital, Rahim Yar Khan

Sheikh Zayed Medical College / Hospital, Rahim Yar Khan
Technical Evaluation Statement (Financial Year 2021-22) on 28-10-2021

ADP Scheme "Programme for Revamping of Emergencies of Tertiary Care Hospital in Punjab (Sheikh Zayed Hospital, Rahim Yar Khan)"

Sr.#	ITEM NO. 2E Hand Sanitizer	
	KNOCK DOWN CRITERIA (COMMERCIAL EVALUATION)	
1	Original Receipt for Purchase of Tender	
2	An Affidavit (as per Specimen Performance Stamp Paper of Rs. 100/-)	
3	Copy of CDR/Bid Security	
4	Bid Validity	
5	Delivery Period	
6	Valid Manufacturing Licence (Embassy Attested)	
	Remarks (Responsive/Non-Responsive)	
Evolution parameter (quoted Items/Equipment)		
1	Exclusive /Sole authorization agreement from the Foreign Manufacturer.	
2	Certificate from the Manufacturer that the after sales Services/backup Services shall be provided jointly with the local sole agent and in case of change of local agent, they will provide the after sale services themselves for through newly appointed agent for the period mentioned from the date of commissioning.	
3	Submission of valid legally enforceable exclusive authorization letter of manufacturer assuring full guarantee and warranty obligations as per enclosed manufacturer authorized form with the bid document.	
4	Valid Manufacturing License of quoted Item Manufacturing embassy attested with certified documentary evidence of manufacturing site & location.	
5	The medical equipment offered from foreign countries of USA, Europe and Japan shall be eligible to participate and must bear FDA510k, CE(MDO) or MHLW (Ministry of Health, Labor and Welfare) standard.	
6	A Certificate from the manufacturer that the installation will be conducted in conformity with the system requirements by following the professional approach.	
7	Satisfactory Past Performance of the bidder for quoted product	
8	The firms shall also declare the make, model, country of origin and country of manufacturing of all standard accessories to be provided with the equipment.	
9	The quoted model of imported product shall be available on the current official website of the manufacturer; otherwise the quoted product shall be considered obsolete/ redundant and will straight away be rejected. (Web link and Screenshot must be attached)	
10	The bid must comply with the required technical specifications of the quoted single item/ complete package. Incomplete offer will straightaway be rejected.	
11	Acceptance of terms and condition of tender documents duly signed and stamped	
	TECHNICAL SPECIFICATIONS	
		Make
		Model
		Country of Manufacturer (Origin): USA, EU, Japan

No Firm Offered this Item

Prof. Dr. Muhammad Saleem
 M.B.S., M.C.P.S., F.C.P.S. (Qual. Medicine)
 Head of Pediatric Department
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan

Prof. Dr. Nuzhat Rashid
 Head of Obs & Gynae Unit-II
 S.Z.M.C. Hospital, Rahim Yar Khan

Dr. Zeeshan Khan
 M.B.B.S., F.C.P.S.
 Assistant Professor Anaesthesia ICU
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan

Eng. AWAIS BIN TARIQ
 Bio Medical Engineer/Technical Officer
 Sheikh Zayed Medical College Hospital P.V.V.

DR. AKIF RAUF
 Deputy Medical Superintendent
 Sheikh Zayed Hospital Rahim Yar Khan

Sheikh Zayed Medical College / Hospital, Rahim Yar Khan
Technical Evaluation Statement (Financial Year 2021-22) on 28-10-2021

ADP Scheme "Programme for Revamping of Emergencies of Tertiary Care Hospital in Punjab (Sheikh Zayed Hospital, Rahim Yar Khan)"

Sr.#	Item.No.03-E Sang Taken black More tube	
	KNOCK DOWN CRITERIA (COMMERCIAL EVALUATION)	
1	Original Receipt for Purchase of Tender	
2	An Affidavit (as per Specimen Performance Stamp Paper of Rs. 100/-)	
3	Copy of CDR/Bid Security	
4	Bid Validity	
5	Delivery Period	
6	Valid Manufacturing Licence (Embassy Attested)	
	Remarks (Responsive/Non-Responsive)	
	Evaluation parameter (quoted items/Equipment)	
1	Exclusive /Sole authorization agreement form the Foreign Manufacturer.	
2	Certificate from the Manufacturer that the after sales Services/backup Services shall be provided jointly with the local sole agent and in case of change of local agent, they will provide the after sale services themselves for through newly appointed agent for the period mentioned from the date of commissioning.	
3	Submission of valid legally enforceable exclusive authorization letter of manufacturer assuring full guarantee and warranty obligations as per enclosed manufacturer authorized form with the bid document.	
4	Valid Manufacturing License of quoted item Manufacturing embassy attested with certified documentary evidence of manufacturing site & location.	
5	The medical equipment offered from foreign countries of USA, Europe and Japan shall be eligible to participate and must bear FDA510k, CE(MDD) or MHLW (Ministry of Health, Labor and Welfare) standard.	
6	A Certificate from the manufacturer that the installation will be conducted in conformity with the system requirements by following the professional approach.	
7	Satisfactory Past Performance of the bidder for quoted product	
8	The firms shall also declare the make, model, country of origin and country of manufacturing of all standard accessories to be provided with the equipment.	
9	The quoted model of imported product shall be available on the current official website of the manufacturer; otherwise the quoted product shall be considered obsolete/ redundant and will straight away be rejected. (Web link and Screenshot must be attached)	
10	The bid must comply with the required technical specifications of the quoted single item/ complete package. Incomplete offer will straightaway be rejected.	
11	Acceptance of terms and condition of tender documents duly signed and stamped	

No Firm Offered this Item

Dr. Muhammad Saleem
 MBBS, MCPS, FCP (Pediatrics)
 Head of Pediatric Department
 Sheikh Zayed Medical College/Hospital
 Rahim Yar Khan

Prof. Dr. Nazhat Raza
 Head of Obs & Gynae Unit-II
 S.Z.M.C/Hospital, Rahim Yar Khan

Dr. Zeeshan Khan
 M.B.B.S., F.C.P.S.
 Assistant Professor Anaesthesia ICU
 Sheikh Zayed Medical College/Hospital
 Rahim Yar Khan

ENG. AWAIS BIN TARIQ
 Bio Medical Engineer/Technical Officer
 Sh. Zayed Medical College/Hospital R.Y.K.

DR. AKIF RAUF
 Deputy Medical Superintendent
 Sheikh Zayed Hospital Rahim Yar Khan

Sheikh Zayed Medical College / Hospital, Rahim Yar Khan
Technical Evaluation Statement (Financial Year 2021-22) on 28-10-2021
ADP Scheme "Programme for Revamping of Emergencies of Tertiary Care Hospital in Punjab (Sheikh Zayed Hospital, Rahim Yar Khan)"

Sr. No.	ITEM NO. 4E O.T Table	M/s Sigma International	M/s Medical Equipment&systems	M/s Mediland Pakistan	M/s Eastern Medical	M/s Vertex Medical	M/s Radiant Medical
	KNOCK DOWN CRITERIA (COMMERCIAL EVALUATION)						
1	Original Receipt for Purchase of Tender	Yes	Yes	Yes	Yes	Yes	Yes
2	An Affidavit (as per Spelman Performa on Stamp Paper of Rs. 100/-)	Yes	Yes	Yes	Yes	Yes	Yes
3	Copy of CDR/Bid Security	Yes	Yes	Yes	Yes	Yes	Yes
4	Bid Validity	Yes	Yes	Yes	Yes	Yes	Yes
5	Delivery Period	Yes	Yes	Yes	Yes	Yes	Yes
6	Valid Manufacturing Licence (Embassy Attested)	Yes (Online Verified Registration Certificate)	Yes (Online Verified Registration Certificate)	Yes (Online Verified Registration Certificate)	Yes (Online Verified Registration Certificate)	Registration certificate not verified online	Yes (Online Verified Registration Certificate)
	Remarks (Responsive/Non-Responsive)	Responsive	Responsive	Responsive	Responsive	Non Responsive	Responsive
	Evaluation parameter (quoted items/Equipment)						
1	Exclusive /Sole authorization agreement form the Foreign Manufacturer.	Yes	Yes	Yes	Yes	Yes	Yes
2	Certificate from the Manufacturer that the after sales Services/backup Services shall be provided jointly with the local sole agent and incase of change of local agent, they will provide the after sale services themselves for through newly appointed agent for the period mentioned from the date of commissioning.	Yes	Yes	Yes	Yes	Yes	Yes
3	Submission of valid legally enforceable exclusive authorization letter of manufacturer assuring full guarantee and warranty (03 Years) obligations as per enclosed manufacturer authorized form with the bid document.	Yes	Yes	Yes	Yes	Yes	Yes
4	Valid Manufacturing License of quoted item Manufacturing embassy attested with certified documentary evidence of manufacturing site & location.	Yes	Yes	Yes	Yes	Registration certificate not verified online	Yes
5	The medical equipment offered from foreign countries of USA, Europe and Japan shall be eligible to participate and must bear FDA510k, CE(MDD) or MHLW (Ministry of Health, Labor and Welfare) standard.	Yes	Yes	Yes	Yes	Yes	Yes
6	A Certificate from the manufacturer that the installation will be conducted in conformity with the system requirements by following the professional approach.	Yes	Yes	Yes	Yes	Yes	Yes
7	Satisfactory Past Performance of the bidder for quoted product	Yes	Yes	Yes	Yes	Yes	No
8	The firms shall also declare the make, model, country of origin and country of manufacturing of all standard accessories to be provided with the equipment.	Yes	Yes	Yes	Yes	Yes	Yes
9	The quoted model of imported product shall be available on the current official website of the manufacturer; otherwise the quoted product shall be considered obsolete/ redundant and will straight away be rejected. (Web link and Screenshot must be attached)	Yes	Yes	Yes	Yes	Yes	Yes

Zeeshan Khan
 M.B.B.S., F.C.P.S.
 Assistant Professor Anaesthesiology (ICU)
 Sheikh Zayed Medical College Hospital

Eng. AWAIS BIN TARIQ
 Bio Medical Engineer/Technical Officer
 Sheikh Zayed Medical College Hospital

DR. AKIF RAUF
 Director
 Sheikh Zayed Medical College Hospital

DR. AKIF RAUF
 Director
 Sheikh Zayed Medical College Hospital

Sr. No.	ITEM NO. 4E O.T Table	M/s Sigma International	M/s Medical Equipment&systems	M/s Mediland Pakistan	M/s Eastern Medical	M/s Vertex Medical	M/s Radiant Medical
10	The bid must comply with the required technical specifications of the quoted single item/ complete package. Incomplete offer will straightaway be rejected.	Yes	Yes	Yes	Yes	Yes	Yes
11	Acceptance of terms and condition of tender documents duly signed and stamped	Yes	Yes	Yes	Yes	Yes	Yes
	Remarks (Responsive/Non-Responsive)	Responsive	Responsive	Responsive	Responsive	Non Responsive	Non Responsive
	Evaluation Criteria (Score)	75	81	83	85	86	79
	TECHNICAL SPECIFICATIONS						
	Make	SU-03(T-shaped base)	Paramount	SCHMITZ	Medifa	ALVO MEDICAL	MEDICINOS
	Model	Famed	KM-401	OPX Mobilis RC30	5000 Series (504220)	ETUDA 4-02	KEPLER-T3000
	Country of Manufacturer (Origin): USA, EU, Japan	Europe	Japan	Germany	Germany	POLAND	LITHUANIA
1	Operating Tables used to conduct the different kind of surgical interventions of patients, the dedicated tables provide all the positions required by the surgeon	Yes	Yes	Yes	Yes	Yes	Yes
2	Weight bearing capacity of 200kg or more	Yes	Yes	Yes	Yes	Yes	Yes
3	4-5 Sectional Operation Table with Single Leg Section	Yes	Yes	Yes	Yes	Yes	Yes
4	Table top equipped with radiolucent material.	Yes	Yes	Yes	Yes	Yes	Yes
5	The mattress covers with washable, antistatic material.	Yes	Yes	Yes	Yes	Yes	Yes
6	X-ray Cassette holder for X-Ray and C-Arm facility	Yes	Yes	Yes	Yes	Yes	Yes
7	Electric Height adjustment: 750 to 1000 mm or more.	Yes	Yes	Yes	No (780 mm)	Yes	Yes
8	Electric Trendelenburg/Reverse Trendelenburg: 25° / -25° or better.	Yes	Yes	Yes	Yes	Yes	Yes
9	Electric lateral tilt: 20° / -20° or better.	Yes	Yes	Yes	Yes	Yes	Yes
10	Manual/electric backrest adjustment: 70° / -15° or better.	Yes	Yes	Yes	Yes	Yes	Yes
11	Manual leg section adjustment: 20° / -90° or better.	Yes	Yes	Yes	Yes	Yes	Yes
12	220-230 V, 50 Hz.	Yes	Yes	Yes	Yes	Yes	Yes
13	Hand control unit.	Yes	Yes	Yes	Yes	Yes	Yes
14	Override panel in the column for back up control in emergency cases.	Yes	Yes	Yes	Yes	Yes	Yes
15	Battery backup control of table in case of main power failure.	Yes	Yes	Yes	Yes	Yes	Yes
	Accessories:						
a	Arm rest with clamp	Yes	Yes	Yes	Yes	Yes	Yes
b	Fixation strap	Yes	Yes	Yes	Yes	Yes	Yes
c	Anesthesia screen	Yes	Yes	Yes	Yes	Yes	Yes
d	Adjustable leg rest pads	Yes	Yes	Yes	Yes	Yes	Yes
e	Large width body strap	Yes	Yes	Yes	Yes	Yes	Yes
f	Adjustable bottle holder rod	Yes	Yes	Yes	Yes	Yes	Yes
g	Shoulder support	Yes	Yes	Yes	Yes	Yes	Yes
	Remarks (Responsive/Non-Responsive)	Responsive	Responsive	Responsive	Non Responsive	Responsive	Responsive

Note:

- 1 M/S VERTEX MEDICAL is Non Responsive as per KNOCK DOWN CRITERIA.
- 2 M/S EASTREN MEDICAL is Non Responsive as per TECHNICAL CRITERIA.
- 3 M/S RADIANT is Non Responsive as per KNOCK DOWN CRITERIA.

Dr. Zeeshan Khan
M.B.B.S., F.C.P.S.
Assistant Professor Anaesthesia /ICU
Sheikh Zayed Medical College/Hospital
Rahim Yar Khan

[Signature]

Eng. AWAIS BIN TARIC
Bio Medical Engineer/Technical Officer
Sh. Zayed Medical College/Hospital

[Signature]

DR. AKIF RAUF
Deputy Medical Superintendent
Sh. Zayed Medical College/Hospital

[Signature]

Sheikh Zayed Medical College / Hospital, Rahim Yar Khan
Technical Evaluation Statement (Financial Year 2021-22) on 28-10-2021

Prof. Dr. Muhammad Ali
 MBBS, MCPS, FICS, FRCGS
 Head of Pediatrics
 Sheikh Zayed Medical College
 Rahim Yar Khan

Page 7 of 29

ADP Scheme "Programme for Revamping of Emergencies of Tertiary Care Hospital in Punjab (Sheikh Zayed Hospital, Rahim Yar Khan)"

Sr. #	ITEM NO. 5 E Pendant (Anesthesia)	M/s Mediland Pakistan	M/s Total Technologies	M/s Vertex Medical
	KNOCK DOWN CRITERIA (COMMERCIAL EVALUATION)			
1	Original Receipt for Purchase of Tender	Yes	Yes	Yes
2	An Affidavit (as per Specimen Performa on Stamp Paper of Rs. 100/-)	Yes	Yes	Yes
3	Copy of CDR/Bid Security	Bank Guarantee	Yes	Yes
4	Bid Validity	Yes	Yes	Yes
5	Delivery Period	Yes	Yes	Yes
6	Valid Manufacturing Licence (Embassy Attested)	Yes (Online Verified Registration Certificate)	Yes (Online Verified Registration Certificate)	Online Not verified as per given website address
	Remarks (Responsive/Non-Responsive)	Responsive	Responsive	Non Responsive
	Evulation parameter (quoted itmes/Equipment)			
1	Exclusive /Sole authorization agreement form the Foreign Manufacturer.	Yes	Yes	Yes
2	Certificate from the Manufacturer that the after sales Services/backup Services shall be provided jointly with the local sole agent and incase of change of local agent, they will provide the after sale services themselves for through newly appointed agent for the period mentioned from the date of commissioning.	Yes	Yes	Yes
3	Submission of valid legally enforceable exclusive authorization letter of manufacturer assuring full guarantee and warranty obligations as per enclosed manufacturer authorized form with the bid document.	Yes	Yes	Yes
4	Valid Manufacturing License of quoted item Manufacturing embassy attested with certified documentary evidence of manufacturing site & location.	Yes	Yes	Online Not verified as per given website address
5	The medical equipment offered from foreign countries of USA, Europe and Japan shall be eligible to participate and must bear FDA510k, CE(MDD) or MHLW (Ministry of Health, Labor and Welfare) standard.	Yes	Yes	Yes
6	A Certificate from the manufacturer that the installation will be conducted in conformity with the system requirements by following the professional approach.	Yes	Yes	Yes
7	Satisfactory Past Performance of the bidder for quoted product	Yes	No	Yes
8	The firms shall also declare the make, model, country of origin and country of manufacturing of all standard accessories to be provided with the equipment.	Yes	Yes	Yes
9	The quoted model of imported product shall be available on the current official website of the manufacturer; otherwise the quoted product shall be considered obsolete/ redundant and will straight away be rejected. (Web link and Screenshot must be attached)	Yes	Yes	Yes
10	The bid must comply with the required technical specifications of the quoted single item/ complete package. Incomplete offer will straightaway be rejected.	Yes	Yes	Yes
11	Acceptance of terms and condition of tender documents duly signed and stamped	Yes	Yes	Yes
	Remarks (Responsive/Non-Responsive)	Responsive	Non Responsive	Non Responsive

Dr. Zeeshan Khan
 MBBS, FRCGS, FICS
 Assistant Professor of Anesthesia
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan

DR. AKIF RAUF
 Deputy Medical Superintendent
 Sheikh Zayed Hospital Rahim Yar Khan

AWAIS BIN TARIQ
 Bio Medical Engineer/Technical Officer
 Sh. Zayed Medical College Hospital R.Y.K.

Prof. Dr. Nuzhat Rashid
 Head of Obs & Gynae Unit-II
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan

Sr. #	ITEM NO. 5 E Pendant (Anesthesia)	M/s Mediland Pakistan	M/s Total Technologies	M/s Vertex Medical
	Evaluation Criteria (Score)	83	84	86
	TECHNICAL SPECIFICATIONS			
	Make	KLS Martin	LM MEDICAL	SORIMA HOSPITAL DEVELOPMENT (SHD)
	Model	Independent Flex	OKI-ISOLA	MOTION
	Country of Manufacturer (Origin): USA, EU, Japan	Germany	Italy	Italy
	Evaluation Criteria Score			
1	Ceiling mounted Anaesthesia Pendant system should be powder coated, hygienically smooth surface, impact and disinfectant resistant	Yes	Yes	Yes
2	Shelves – 3 or 4Pcs	Yes	Yes	Yes
3	Drawer – 1 or 2 Pcs	Yes	Yes	Yes
4	No. of Arms: 2	Yes	Yes	Yes
5	Length of first arm: 1000mm	Yes	Yes	Yes
6	Length of second arm: 800mm	Yes	Yes	Yes
7	Max. Degree of rotation at each arm: 300° or better	Yes	Yes	Yes
8	Max. Loading capacity: 200Kg or better	Yes	Yes	Yes
9	Braking system: Electromagnetic / Electric / pneumatic brakes to control horizontal movement	Yes	Yes	Yes
10	Electrical Sockets British Standard – 16 to 20 Pcs	Yes	Yes	Yes
11	LAN Data ports RJ45 for PC – 1 Pc	Yes	Yes	Yes
12	Telephone Outlets RJ11 – 1Pc	Yes	Yes	Yes
13	Oxygen Outlet: 2Pcs	Yes	Yes	Yes
14	Medical Air 4 Bar: 2Pcs	Yes	Yes	Yes
15	Vacuum Outlet: 2Pcs	Yes	Yes	Yes
16	Nitrous Oxide Outlet: 2Pcs	Yes	Yes	Yes
17	AGSS Outlet: 1Pc	Yes	Yes	Yes
18	I.V Pole – 1 Pc	Yes	Yes	Yes
19	Basket for accessories	Yes	Yes	Yes
	Remarks (Responsive/Non-Responsive)	Responsive	Responsive	Responsive

Note:

- 1 M/S TOTAL TECHNOLOGIES is Non Responsive as per KNOCK DOWN CRITERIA.
- 2 M/S VERTEX MEDICAL is Non Responsive as per KNOCK DOWN CRITERIA.

Dr. Zeeshan Khan

M.B.B.S., F.C.P.S.
Assistant Professor Anaesthesia / ICU
Sh. Zayed Medical College Hospital
Rahim Yar Khan

DR. AKIF RAUF

Deputy Medical Superintendent
Sh. Zayed Medical College Hospital
Rahim Yar Khan

Eng. AWAIS BIN TARIQ
Bio Medical Engineer/Technical Officer
Sh. Zayed Medical College Hospital R.Y.K.

Dr. Muhammad Iqbal
M.B.B.S., F.C.P.S. (Anesthesia)
Assistant Professor Anaesthesia / ICU
Sh. Zayed Medical College Hospital
Rahim Yar Khan

Prof. Dr. Muhammad Iqbal
Head of Obs & Gynae
S.Z.M.C Hospital, Rahim Yar Khan

Sheikh Zayed Medical College / Hospital, Rahim Yar Khan
Technical Evaluation Statement (Financial Year 2021-22) on 28-10-2021

ADP Scheme "Programme for Revamping of Emergencies of Tertiary Care Hospital in Punjab (Sheikh Zayed Hospital, Rahim Yar Khan)

Sr.#	ITEM NO. 6 E Pendant (Surgical)	M/s Mediland Pakistan	M/s Total Technologies	M/s Verity Medical
	KNOCK DOWN CRITERIA (COMMERCIAL EVALUATION)			
1	Original Receipt for Purchase of Tender	Yes	Yes	Yes
2	An Affidavit (as per Specimen Performance Stamp Paper of Rs. 100/-)	Yes	Yes	Yes
3	Copy of CDR/Bid Security	Bank Guarantee	Yes	Yes
4	Bid Validity	Yes	Yes	Yes
5	Delivery Period	Yes	Yes	Yes
6	Valid Manufacturing Licence (Embassy Attested)	Yes (Online Verified Registration Certificate)	Yes (Online Verified Registration Certificate)	Online Not verified registration certificate
	Remarks (Responsive/Non-Responsive)	Responsive	Responsive	Non Responsive
	Evaluation parameter (quoted items/Equipment)			
1	Exclusive /Sole authorization agreement from the Foreign Manufacturer.	Yes	Yes	Yes
2	Certificate from the Manufacturer that the after sales Services/backup Services shall be provided jointly with the local sole agent and in case of change of local agent, they will provide the after sale services themselves for through newly appointed agent for the period mentioned from the date of commissioning.	Yes	Yes	Yes
3	Submission of valid legally enforceable exclusive authorization letter of manufacturer assuring full guarantee and warranty obligations as per enclosed manufacturer authorized form with the bid document.	Yes	Yes	Yes
4	Valid Manufacturing License of quoted Item Manufacturing embassy attested with certified documentary evidence of manufacturing site & location.	Yes	Yes	Online Not verified registration certificate
5	The medical equipment offered from foreign countries of USA, Europe and Japan shall be eligible to participate and must bear FDA510k, CE(MDD) or MHLW (Ministry of Health, Labor and Welfare) standard.	Yes	Yes	Yes
6	A Certificate from the manufacturer that the installation will be conducted in conformity with the system requirements by following the professional approach.	Yes	Yes	Yes
7	Satisfactory Past Performance of the bidder for quoted product.	Yes	No	No
8	The firms shall also declare the make, model, country of origin and country of manufacturing of all standard accessories to be provided with the equipment.	Yes	Yes	Yes
9	The quoted model of imported product shall be available on the current official website of the manufacturer; otherwise the quoted product shall be considered obsolete/ redundant and will straight away be rejected. (Web link and Screenshot must be attached)	Yes	Yes	Yes
10	The bid must comply with the required technical specifications of the quoted single item/ complete package. Incomplete offer will straightaway be rejected.	Yes	Yes	Yes
11	Acceptance of terms and condition of tender documents duly signed and stamped	Yes	Yes	Yes
		Responsive	Non Responsive	Non Responsive

Dr. Zeeshan Khan

M.B.S., F.C.P.S.
 Anaesthesia / ICU
 College Hospital
 Rahim Yar Khan

DR. AKIE RAUF

Deputy Medical Superintendent
 Sheikh Zayed Hospital, Rahim Yar Khan

ENG. AWAZ BIN TARIQ

Bio Medical Engineer/Technical Officer
 Sheikh Zayed Hospital, Rahim Yar Khan

Prof. Dr. Nuzhat Kashif

Head of Obs & Gynae Unit-II
 S.Z.M.C/Hospital, Rahim Yar Khan

Sr.#	ITEM NO. 6 E Pendant (Surgical)	M/s Mediland Pakistan	M/s Total Technologies	M/s Vertex Medical
	TECHNICAL SPECIFICATIONS			
	Make	KLS Martin	LM MEDICAL	SORIMA HOSPITAL DEVELOPMENT (SHD)
	Model	Independent Flex	OKI-ISOLA	MOTION
	Country of Manufacturer (Origin): USA, EU, Japan	Germany	Italy	Italy
	Evaluation Criteria Score	83	84	86
1	Ceiling mounted Surgical Pendant system should be powder coated, hygienically smooth surface, impact and disinfectant resistant	Yes	Yes	Yes
2	Shelves – 3 or 4Pcs	Yes	Yes	Yes
3	Drawer – 1 or 2 Pcs	Yes	Yes	Yes
4	No. of Arms: 2	Yes	Yes	Yes
5	Length of first arm: 1000mm	Yes	Yes	Yes
6	Length of second arm: 800mm	Yes	Yes	Yes
7	Max. Degree of rotation at each arm: 300° or better	Yes	Yes	Yes
8	Max. Loading capacity: 200Kg or better	Yes	Yes	Yes
9	Braking system: Electromagnetic / Electric / pneumatic brakes to control horizontal movement	Yes	Yes	Yes
10	Electrical Sockets British Standard – 16 to 20 Pcs	Yes	Yes	Yes
11	LAN Data ports RJ45 for PC – 1 Pc	Yes	Yes	Yes
12	Telephone Outlets RJ11 – 1Pc	Yes	Yes	Yes
13	Oxygen Outlet: 2Pcs	Yes	Yes	Yes
14	Medical Air 4 Bar: 2Pcs	Yes	Yes	Yes
15	Vacuum Outlet: 2Pcs	Yes	Yes	Yes
16	I.V Pole – 1 Pc	Yes	Yes	Yes
		Responsive	Responsive	Responsive

Note:

- 1 M/S VERTEX MEDICAL is Non Responsive as per KNOCK DOWN CRITERIA.
- 2 M/S TOTAL TECHNOLOGIES is Non Responsive as per KNOCK DOWN CRITERIA.

Dr. Zeeshan Khan
M.B.B.S., F.C.P.S.
Assistant Professor Anaesthesia/ICU
Sheikh Zayed Medical College Hospital,
Rahim Yar Khan

DR. AKIF RAUF
Deputy Medical Superintendent
Sheikh Zayed Medical College Hospital, Rahim Yar Khan

Eng. AWAIS BIN TARIQ
Bio Medical Engineer/Technical Officer
Sh. Zayed Medical College Hospital R YK

Prof. Dr. Muhammad Saleem
M.B.B.S., F.C.P.S., F.R.C.S.
Head of Pediatric Department
Sheikh Zayed Medical College Hospital,
Rahim Yar Khan

Dr. Muhammad Zafar Khan
M.B.B.S., F.C.P.S., F.R.C.S.
Assistant Professor Anaesthesia/ICU
Sheikh Zayed Medical College Hospital,
Rahim Yar Khan

Dr. Zahid Raheem
M.B.B.S., F.C.P.S., F.R.C.S.
Head of OHS & Gynaecology
Sh. Zayed Medical College Hospital, Rahim Yar Khan

Sheikh Zayed Medical College / Hospital, Rahim Yar Khan
Technical Evaluation Statement (Financial Year 2021-22) on 28-10-2021

ADP Scheme "Programme for Revamping of Emergencies of Tertiary Care Hospital in Punjab (Sheikh Zayed Hospital, Rahim Yar Khan)

Sr. #.	ITEM NO. 7 E HIP Spica O.T Table	M/s Ashraf Nadeem
	KNOCK DOWN CRITERIA (COMMERCIAL EVALUATION)	
1	Original Receipt for Purchase of Tender	Yes
2	An Affidavit (as per Speimen Performa on Stamp Paper of Rs. 100/-)	Yes
3	Copy of CDR/Bid Security	Yes
4	Bid Validity	Yes
5	Delivery Period	Yes
6	Valid Manufacturing Licence (Embassy Attested)	No
	Remarks (Responsive/Non-Responsive)	Non Responsive
	Evulation parameter (quoted Itmes/Equipment)	
1	Exclusive /Sole authorization agreement form the Foreign Manufacturer.	No
2	Certificate from the Manufacturer that the after sales Services/backup Services shall be provided jointly with the local sole agent and incase of change of local agent, they will provide the after sale services themself for through newly oppointed agent for the period mentioned from the date of commissioning.	No
3	Submission of valid legally enforceable exclusive authorization letter of manufacturer assuring full guarantee and warranty obligations as per enclosed manufacturer authorized form with the bid document.	No
4	Valid Manufacturing License of quoted Item Manufacturing embassy attested with certified documentary evidence of manufacturing site & location.	No
5	The medical equipment offered from foreign countries of USA, Europe and Japan shall be eligible to participate and must bear FDA510k, CE(MDD) or MHLW (Ministry of Health, Labor and Welfare) standard.	No
6	A Certificate from the manufacturer that the installation will be conducted in conformity with the system requirements by following the professional approach.	No
7	Satisfactory Past Performance of the bidder for quoted product	No
8	The firms shall also declare the make, model, country of origin and country of manufacturing of all standard accessories to be provided with the equipment.	No
9	The quoted model of imported product shall be available on the current official website of the manufacturer; otherwise the quoted product shall be considered obsolete/ redundant and will straight away be rejected. (Web link and Screenshot must be attached)	No
10	The bid must comply with the required technical specifications of the quoted single item/ complete package. Incomplete offer will straightaway be rejected.	No
11	Acceptance of terms and condition of tender documents duly signed and stamped	No
	Remarks (Responsive/Non-Responsive)	Non Responsive
	TECHNICAL SPECIFICATIONS	
	Make	RAS SURGICAL
	Model	Local
	Country of Manufacturer (Origin): USA, EU, Japan	Local

Prof. Dr. Muhammad Ashraf Nadeem
 M.D., M.C.P.S., F.C.P.S. (Gen Med) - MRCP
 Head of Pediatric Department
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan

Prof. Dr. Anzhat Rashid
 Head of Obs & Gynae Unit-II
 S.Z.M. CH Hospital, Rahim Yar Khan

Dr. Zeeshan Khan
 M.B.B.S., F.C.S., F.A.S.
 Assistant Professor Anaesthesia ICU
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan

Eng. AWAIS BIN IARIG
 Bio Medical Engineer/Technical Officer
 Sh. Zayed Medical College Hospital

DR. AKIF RAUF
 Deputy Medical Superintendent
 Sheikh Zayed Hospital Rahim Yar Khan

Sr. #.	ITEM NO. 7 E HIP Spica O.T Table	M/s Ashraf Nadeem
	The table and its system storing, study enough to hold weight of 225 kg or more 5 sectional operation table with single leg function.	No
	Table top equipped with radiolucent material.	No
	Width 540 mm or more, length 2000 mm or more	No
	The mattress covers with washable, antistatic material.	No
	X-Ray cassett tunnel throughout the table.	No
	single Leg Sections	No
	Fulfils X rays and C arm requiremnts.	No
	Following motor driven movements controlled through hand set controller.	No
	Height adjustment: 595 to 890 mm or more.	No
	Trendelenburg / Reverse Trendelenburg: 26° / -26° or better	No
	Backrest adjustment 70 Deg. / -15 Deg. Or better	No
	Leg section adjustment motor driven: 20 deg. / -90 deg. Or better	No
	Lateral tilt 20 deg. / -20 deg.	No
	220-220 V, connction	No
	hand control Unit	No
	Wireless / IR Hand control Unit	No
	Override panel in the column for Back up control in emergency cases	No
	Battery backup control of table in case of main power failure.	No
	Accessories:	-
	Anesthesia Screen (01 No.)	No
	Arm Rest with straes (01 pair)	No
	IV Rod (01 No.)	No
	Devon strap (Body strap) Large width (01 No.)	No
	Leg and Body strap	No
	Orthopedic Extension Device	No
	Extension Device Height: 700 - 1160 mm	No
	Total Length 1857	No
	Pelvic support pad angle +/- 20 Degree and +/-40 Degree.	No
	Perineal post pad	No
	Traction device.	No
	Stainless steel Accessory stand.	No
	Remarks (Responsive/Non-Responsive)	Non Responsive

Note:

M/S Ashraf Nadeem is Non Responsive as per KNOCK DOWN CRITERIA.

Zeeshan
Dr. Zeeshan Khan
M.D.B.S., F.C.P.S.,
Assistant Professor, Aga Khan Hospital
Sheikh Zayed Medical College Hospital
Rahim Yar Khan

Akif Rauf
DR. AKIF RAUF
Deputy Medical Superintendent
Sheikh Zayed Hospital Rahim Yar Khan

Eng. Awaiz Bin Tariq
Eng. AWAIZ BIN TARIQ
Bio Medical Engineer Technical Officer
Sh. Zayed Medical College Hospital R.Y.K.

Prof. Dr. Nuzhat Rasheed
Prof. Dr. Nuzhat Rasheed
Head of Obs & Gynae Unit-II
S.Z.M.C/Hospital, Rahim Yar Khan

Prof. Dr. Muhammad Saleem
Prof. Dr. Muhammad Saleem
M.D.S., F.C.P.S., (Gold Medalist)
Head of Pediatric Department
Sheikh Zayed Medical College Hospital
Rahim Yar Khan

Sheikh Zayed Medical College / Hospital, Rahim Yar Khan
Technical Evaluation Statement (Financial Year 2021-22) on 28-10-2021

ADP Scheme "Programme for Revamping of Emergencies of Tertiary Care Hospital in Punjab (Sheikh Zayed Hospital, Rahim Yar Khan)

Sr.#	ITEM NO. 8E-Medium Steam Sterilizer	M/s Sigma International	M/S Global	M/S Clinical Life	M/s Mediland Pakistan	M/s Total Technologies	M/s Vertex Medical
	KNOCK DOWN CRITERIA (COMMERCIAL EVALUATION)						
1	Original Receipt for Purchase of Tender	Yes	Yes	Yes	Yes	Yes	Yes
2	An Affidavit (as per Specimen Performance Stamp Paper of Rs. 100/-)	Yes	Yes	Yes	Yes	Yes	Yes
3	Copy of CDR/Bid Security	Yes	Yes	Yes	Bank Guarantee	Yes	Yes
4	Bid Validity	Yes	Yes	Yes	Yes	Yes	Yes
5	Delivery Period	Yes	Yes	Yes	Yes	Yes	Yes
6	Valid Manufacturing Licence (Embassy Attested)	Yes (Online Verified Registration Certificate)	Not Provided	Yes (Online Verified Registration Certificate)	Yes (Online Verified Registration Certificate)	Yes (Online Verified Registration Certificate)	Model not mentioned in Provided Ministry Letter & also not provided online verification of Registration Link
	Remarks (Responsive/Non-Responsive)	Responsive	Non Responsive	Responsive	Responsive	Responsive	Non Responsive
	Evaluation parameter (quoted items/Equipment)						
1	Exclusive /Sole authorization agreement from the Foreign Manufacturer.	Yes	Yes	Yes	Yes	Yes	Yes
2	Certificate from the Manufacturer that the after sales Services/backup Services shall be provided jointly with the local sole agent and in case of change of local agent, they will provide the after sale services themselves for through newly appointed agent for the period mentioned from the date of commissioning.	Yes	Yes	Yes	Yes	Yes	Yes
3	Submission of valid legally enforceable exclusive authorization letter of manufacturer assuring full guarantee and warranty obligations as per enclosed manufacturer authorized form with the bid document.	Yes	Yes	Yes	Yes	Yes	Yes
4	Valid Manufacturing License of quoted item Manufacturing embassy attested with certified documentary evidence of manufacturing site & location.	Yes (Online Verified Registration Certificate)	Not Provided	Yes (Online Verified Registration Certificate)	Yes (Online Verified Registration Certificate)	Yes (Online Verified Registration Certificate)	Model not mentioned in Provided Ministry Letter & also not provided online verification of Registration Link
5	The medical equipment offered from foreign countries of USA, Europe and Japan shall be eligible to participate and must bear FDA510k, CE(MDD) or MHLW (Ministry of Health, Labor and Welfare) standard.	Yes	Yes	Yes	Yes	Yes	Yes
6	A Certificate from the manufacturer that the installation will be conducted in conformity with the system requirements by following the professional approach.	Yes	Yes	Yes	Yes	Yes	Yes
7	Satisfactory Past Performance of the bidder for quoted product. (02 Years)	Yes	Yes	Yes	Yes	Yes	Yes
8	The firms shall also declare the make, model, country of origin and country of manufacturing of all standard accessories to be provided with the equipment.	Yes	Yes	Yes	Yes	Yes	Yes

Dr. Zeeshan Khan
M.B.B.S., F.C.P.S.
Assistant Professor Anaesthesia /ICU
Sheikh Zayed Medical College/Hospital
Rahim Yar Khan

DR. AKIF RAUF
Deputy Medical Superintendent
Sheikh Zayed Hospital Rahim Yar Khan

Eng. AWAIS BIN IARIC
Bio Medical Engineer/Technical Officer
Sheikh Zayed Medical College/Hospital

Prof. Dr. M. M. Chaudhary
Head of OIC
Z.M.C. Hospital, Rahim Yar Khan

Prof. Dr. M. M. Chaudhary
Head of OIC
Z.M.C. Hospital, Rahim Yar Khan

Prof. Dr. M. M. Chaudhary
Head of OIC
Z.M.C. Hospital, Rahim Yar Khan

Sr.#	ITEM NO. 8E-Medium Steam Sterilizer	M/s Sigma International	M/S Global	M/S Clinical Life	M/s Medland Pakistan	M/s Total Technologies	M/s Vertex Medical
9	The quoted model of imported product shall be available on the current official website of the manufacturer; otherwise the quoted product shall be considered obsolete/ redundant and will straight away be rejected. (Web link and Screenshot must be attached)	Yes	Yes	Yes	Yes	Yes	Yes
10	The bid must comply with the required technical specifications of the quoted single item/ complete package. Incomplete offer will straightaway be rejected.	Yes	Yes	Yes	Yes	Yes	Yes
11	Acceptance of terms and condition of tender documents duly signed and stamped	Yes	Yes	Yes	Yes	Yes	Yes
	Remarks (Responsive/Non-Responsive)	Responsive	Non-Responsive	Responsive	Responsive	Responsive	Non-Responsive
	Evaluation Criteria Score	75	73	77	83	84	81
	TECHNICAL SPECIFICATIONS						
	Make	Celitron	Steris Corporatiton	A.J.Costa	Getinge Infection c	MMM Group	STEELCO, SPA
	Model	A-666	AMSCO 400	Amaro 5000 707070-2D GI	GSS67H10	Selectomatg PL 666-2 CL	VS 4/2 E/V
	Country of Manufacturer (Origin): USA, EU, Japan	Europe	USA	Portugal	Poland	Germany	ITALY
1	High pressure Steam Sterilizer each with external steam supply as primary source with integrated steam generator as backup	Yes	Yes	Yes	Yes	Yes	Yes
2	Fully automatic, programmable, microprocessor type	Yes	Yes	Yes	Yes	Yes	Yes
3	Touch screen colored display and integrated printer	Yes	Yes	Yes	Yes	Yes	Yes
4	Automatic motorized/pneumatic two doors pass through system	Yes	Yes	Yes	Yes	Yes	Yes
5	Time cycled, working pressure 32 psi	Yes	Yes	Yes	Yes	Yes	Yes
6	Safety interlock. Temperature & Pressure recorder. Chamber pressure indicator	Yes	Yes	Yes	Yes	Yes	Yes
7	Cycle indicator to determine the phase of sterilization cycle	Yes	Yes	Yes	Yes	Yes	Yes
8	Program/Cycle selection	Yes	Yes	Yes	Yes	Yes	Yes
9	Complete with standard accessories and removable shelves, capable of taking both packets and containers of all standard sizes	Yes	Yes	Yes	Yes	Yes	Yes
10	Chamber capacity 04 STU, rectangular shape	Yes	Yes	Yes	Yes	Yes	Yes
11	Chamber, jacket and doors made of AISI 316 L/TI	Yes	Yes	Yes	Yes	Yes	Yes
12	The system complete with built-in water saving system, automatic heat exchanger and Air detector	Yes	Yes	Yes	Yes	Yes	Yes
13	Two loading / unloading trolleys and one loading carts compatible with system	Yes	Yes	Yes	Yes	Yes	Yes
14	UPS of suitable capacity with minimum 15 minutes for Controller cum display for monitoring and controlling of parameters during power shedding provided/ installed by the manufacturer	Yes	Yes	Yes	Yes	Yes	Yes
15	Optional: R.O Plant should be quoted	-	-	-	-	-	-
a	R.O Plant 50 gallon per day as a part of standard specification (Local Made)	Yes	Yes	Yes	Yes	Yes	Yes
b	The firm should quote the price of RO plant with the sterilizer.	Yes	Yes	Yes	Yes	Yes	Yes
	Remarks (Responsive/Non-Responsive)	Responsive	Responsive	Responsive	Responsive	Responsive	Responsive

Note:

- 1 M/S GLOBAL is Non Responsive as per KNOCK DOWN CRITERIA.
- 2 M/S VERTEX MEDICAL is Non Responsive as per KNOCK DOWN CRITERIA.

Dr. Zeeshan Khan
M.B.B.S., F.C.P.S.
Assistant Professor Anaesthesia/ICU
Sheikh Zayed Medical College/Hospital
Rahim Yar Khan

DR. AKIF RAUF
Deputy Medical Superintendent
Sheikh Zayed Hospital Rahim Yar Khan

Eng. AWAN BIN IARIC
Bio Medical Engineer/Technical Officer
Sh Zayed Medical College/Hospital

Prof. Dr. M. N. Hashmi
Head of Obs & Gynaecology
Rahim Yar Khan

Sheikh Zayed Medical College / Hospital, Rahim Yar Khan
Technical Evaluation Statement (Financial Year 2021-22) on 28-10-2021

ADP Scheme "Programme for Revamping of Emergencies of Tertiary Care Hospital in Punjab (Sheikh Zayed Hospital, Rahim Yar Khan)

Sr.#	ITEM NO. 9E-RO Plant	M/S Hospital Supply Corporation	M/s Ashraf Nadeem
	KNOCK DOWN CRITERIA (COMMERCIAL EVALUATION)		
1	Original Receipt for Purchase of Tender	Yes	Yes
2	An Affidavit (as per Specimen Performance Stamp Paper of Rs. 100/-)	Yes	Yes
3	Copy of CDR/Bid Security	Yes	Yes
4	Bid Validity	Yes	Yes
5	Delivery Period	Yes	Yes
6	Valid Manufacturing Licence (Embassy Attested)	Local	Local
	Remarks (Responsive/Non-Responsive)	Responsive	Responsive
	Evaluation parameter (quoted items/Equipment)		
1	Exclusive /Sole authorization agreement from the Foreign Manufacturer.	Local	Local
2	Certificate from the Manufacturer that the after sales Services/backup Services shall be provided jointly with the local sole agent and in case of change of local agent, they will provide the after sale services themselves for through newly appointed agent for the period mentioned from the date of commissioning.	Local	Local
3	Submission of valid legally enforceable exclusive authorization letter of manufacturer assuring full guarantee and warranty obligations as per enclosed manufacturer authorized form with the bid document.	Local	Local
4	Valid Manufacturing License of quoted item Manufacturing embassy attested with certified documentary evidence of manufacturing site & location.	Local	Local
5	The medical equipment offered from foreign countries of USA, Europe and Japan shall be eligible to participate and must bear FDA510k, CE(MDD) or MHLW (Ministry of Health, Labor and Welfare) standard.	Local	Local
6	A Certificate from the manufacturer that the installation will be conducted in conformity with the system requirements by following the professional approach.	Local	Local
7	Satisfactory Past Performance of the bidder for quoted product. (02 Years)	Yes	Yes
8	The firms shall also declare the make, model, country of origin and country of manufacturing of all standard accessories to be provided with the equipment.	Local	Local
9	The quoted model of imported product shall be available on the current official website of the manufacturer; otherwise the quoted product shall be considered obsolete/ redundant and will straight away be rejected. (Web link and Screenshot must be attached)	Local	Local
10	The bid must comply with the required technical specifications of the quoted single item/ complete package. Incomplete offer will straightaway be rejected.	Local	Local
11	Acceptance of terms and condition of tender documents duly signed and stamped.	Local	Local
	Remarks (Responsive/Non-Responsive)	Responsive	Responsive
	Technical Specification		
	Make	AQUAMED SOLUTIONS	Hydranautics
	Model	AMS-15000GPD	220-LPH

Prof. Dr. Muhammad Saleem
 M.B.B.S., F.C.P.S. (Gold Medalist)
 Head of Pediatric Department
 Sheikh Zayed Medical College/Hospital
 Rahim Yar Khan

Prof. Dr. Muhtasim Rashid
 Head of Obs & Gynae Unit-II
 S.Z.M.C.Hospital, Rahim Yar Khan

Dr. Zeeshan Khan
 M.B.B.S., F.C.P.S.
 Assistant Professor Anaesthesia/ICU
 Sheikh Zayed Medical College/Hospital
 Rahim Yar Khan

Eng. AWAIS BIN TARIQ
 Medical Engineer/Technical
 Head of Engineering
 Sheikh Zayed Medical College/Hospital
 Rahim Yar Khan

DR. AKIF RAUF
 Deputy Medical Superintendent
 Sheikh Zayed Medical College/Hospital
 Rahim Yar Khan

Sr.#	ITEM NO. 9E-RO Plant	M/S Hospital Supply Corporation	M/s Ashraf Nadeem
	Country of Manufacturer (Origin): USA, EU, Japan	Local	USA
	Evaluation Criteria Score	82	71
1	Including Civil/Plumbing Work and Installation	Yes	Yes
2	RO system should be compatible with the CSSD equipment requirement	Yes	Yes
3	in accordance with the quality of the local water where it is being installed	Yes	Yes
4	It should have imported parts that may be locally assembled	Yes	Yes
5	The reverse osmosis unit including filters, hardness stabilizing unit, activated carbon with prefilter 5um	Yes	Yes
6	Pre-filter 1um, DS reverse osmosis system. Capacity of 150 or higher L / h	Yes	Yes
7	Storage tank 1,000L (for laundry and CSSD).	Yes	Yes
8	Level control for tank, vent filter, pressure pump, ion exchanger, conductivity meter control, distribution manifold, hose set and installation material	Yes	Yes
9	Total Dissolved Salts Level (TDS) of Filtered water should not be more than 15 PPM.	Yes	Yes
	Remarks (Responsive/Non-Responsive)	Responsive	Responsive

Zeshan
Dr. Zeeshan Khan
 M.B.B.S., F.C.P.S.
 Assistant Professor Anaesthesia ICU
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan

[Signature]
Dr. Tariq
 Head of Obs & Gynae Unit-II
 S.Z.M.C Hospital, Rahim Yar Khan

[Signature]
DR. ARIF RAUF
 Deputy Medical Superintendent
 Sheikh Zayed Hospital Rahim Yar Kha

[Signature]
Eng. AWAIS BIN TARIQ
 Bio Medical Engineer/Technical Officer
 Sh. Zayed Medical College Hospital R.Y.K

[Signature]
Dr. Ashraf Nadeem
 M.D.S., F.C.P.S. (Med) F.C.C.P. (Cardiology)
 Assistant Professor Cardiology
 Sheikh Zayed Medical College Hospital R.Y.K

[Signature]
Prof. Dr. Muhammad Saleem
 M.D.S., M.C.S., F.C.P.S. (Gold Medalist)
 Head of Pediatric Department
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan

[Signature]
Prof. Dr. Muhammad Saleem
 M.D.S., M.C.S., F.C.P.S. (Gold Medalist)
 Head of Pediatric Department
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan

Sheikh Zayed Medical College / Hospital, Rahim Yar Khan
Technical Evaluation Statement (Financial Year 2021-22) on 28-10-2021

ADP Scheme "Programme for Revamping of Emergencies of Tertiary Care Hospital in Punjab (Sheikh Zayed Hospital, Rahim Yar Khan)

Sr.#	ITEM NO. 10E-Vascular Set (Local)	M/s Ashraf Nadeem
	KNOCK DOWN CRITERIA (COMMERCIAL EVALUATION)	
1	Original Receipt for Purchase of Tender	Yes
2	An Affidavit (as per Specimen Performance Stamp Paper of Rs. 100/-)	Yes
3	Copy of CDR/Bid Security	Yes
4	Bid Validity	Yes
5	Delivery Period	Yes
6	Valid Manufacturing Licence (Embassy Attested)	Local
	Remarks (Responsive/Non-Responsive)	Responsive
Evaluation parameter (quoted items/Equipment)		
1	Exclusive /Sole authorization agreement from the Foreign Manufacturer.	Local
2	Certificate from the Manufacturer that the after sales Services/backup Services shall be provided jointly with the local sole agent and incase of change of local agent, they will provide the after sale services themselves for through newly appointed agent for the period mentioned from the date of commissioning.	Local
3	Submission of valid legally enforceable exclusive authorization letter of manufacturer assuring full guarantee and warranty obligations as per enclosed manufacturer authorized form with the bid document.	Local
4	Valid Manufacturing License of quoted item Manufacturing embassy attested with certified documentary evidence of manufacturing site & location.	Local
5	The medical equipment offered from foreign countries of USA, Europe and Japan shall be eligible to participate and must bear FDA510k, CE(MDD) or MHLW (Ministry of Health, Labor and Welfare) standard.	Local
6	A Certificate from the manufacturer that the installation will be conducted in conformity with the system requirements by following the professional approach.	Local
7	Satisfactory Past Performance of the bidder for quoted product	Local
8	The firms shall also declare the make, model, country of origin and country of manufacturing of all standard accessories to be provided with the equipment.	Local
9	The quoted model of imported product shall be available on the current official website of the manufacturer; otherwise the quoted product shall be considered obsolete/ redundant and will straight away be rejected. (Web link and Screenshot must be attached)	Local
10	The bid must comply with the required technical specifications of the quoted single item/ complete package. Incomplete offer will straightaway be rejected.	Local
11	Acceptance of terms and condition of tender documents duly signed and stamped	Local
	Remarks (Responsive/Non-Responsive)	Responsive
	Evaluation Criteria Score	71

Prof. Dr. Muhammad Saleem
 M.D., M.P.S. (Child Medicine)
 Head of Pediatric Department
 Sheikh Zayed Medical College/Hospital
 Rahim Yar Khan

Prof. Dr. Zahid Rashid
 Head of Obs & Gynae Unit-II
 S.Z.M.C/Hospital, Rahim Yar Khan

Dr. Zia Khan
 M.B.B.S., F.C.S.
 Assistant Professor, Paediatrics ICU
 Sheikh Zayed Medical College/Hospital
 Rahim Yar Khan

Eng. AWAIS BILAL
 M.B.E., F.C.T.
 Head of Engineering Department
 Sheikh Zayed Medical College/Hospital
 Rahim Yar Khan

DR. AKIF RAUF
 Deputy Medical Superintendent
 Sheikh Zayed Hospital Rahim Yar Khan

Sr.#	ITEM NO. 10E-Vascular Set (Local)	M/s Ashraf Nadeem
	Technical Specifications	
1	POTTS-DE MARTEL SCISSORS25DGS/S185MM	Yes
2	POTTS-DE MARTEL SCISSORS45DGS/S185MM	Yes
3	WEITLANER RETRACTOR 2X3T.SH.105MM	Yes
4	WEITLANER RETRACTOR 3X4T.SH.130MM	Yes
5	WEITLANER RETRACTOR 3X4T.SH.165MM	Yes
6	TC RYDER NEEDLE HOLDERDELSERR 135MM	Yes
7	TC RYDER NEEDLE HOLDERX-DELSERR175MM	Yes
8	TC RYDER NEEDLE HOLDERDELSERR 195MM	Yes
9	DUROGRIP NDL HLDR SERR 0.2/180MM	Yes
10	FERG-FRAZIER SUCT 6FR 2/110MM WRK-LGTH	Yes
11	DE'BAKEY ATR.BULLDOGCLAMP STR.30/90MM	Yes
12	DE'BAKEY ATR.BULLDOGCLAMP STR.45/105MM	Yes
13	DE'BAKEY ATR.BULLDOGCLAMP STR.65/125MM	Yes
14	PEAN-DE'BAKEY ATR.FRCPS STR155MM	Yes
15	DE'BAKEY-PEAN ATR.FRCPS STR.60MM 200MM	Yes
16	DE'BAKEY-PEAN ATR.FCPS STR89MM265MM	Yes
17	DE'BAKEY AORTA ANEURYSM CLMPCVD270MM	Yes
18	DE'BAKEY AORTA ANEURYSM CLAMPCVD330MM	Yes
19	DELICATE THUMB FORCEPS SMOOTH 200MM	Yes
20	STANDARD FORCEPS SERR 145MM	Yes
21	STANDARD FORCEPS SERR 200MM	Yes
22	METZENBAUM SCISSORS CVD 145MM	Yes
23	METZENBAUM SCISSORS CVD 200MM	Yes
24	UNIVERSAL SCISSORS BLACK 145MM	Yes
25	UNIVERSAL SCISSORS BLACK 180MM	Yes
26	ROCHESTER-PEAN FORCEPS STR 160MM	Yes
27	ROCHESTER-PEAN FORCEPS CVD 160MM	Yes
28	HALSTED-MOSQUITO FORCEPS DELSTR125MM	Yes
29	HALSTED-MOSQUITO FORCEPS DEL CVD125MM	Yes
30	DEAVER RETRACTOR FIG 1 25MM305MM	Yes
31	DEAVER RETRACTOR FIG 2 25MM 318MM	Yes
32	DEAVER RETRACTOR FIG 3 38MM305MM	Yes
33	DEAVER RETRACTOR FIG 4 50MM 311MM	Yes
34	DEAVER RETRACTOR FIG 5 75MM 324MM	Yes
35	FINOCHIETTO RIB SPREADER 36X45MM BLDS	Yes
36	LANGENBECK RETRACTOR 40X13MM 210MM	Yes
37	ALLISON LUNG SPATULA 40MM253MM	Yes
38	ALLISON LUNG SPATULA 54MM 324MM	Yes
39	SCALPEL HANDLE #3	Yes
40	SCALPEL HANDLE #4	Yes
41	SCALPEL HANDLE #3L	Yes
42	POTTS-DE MARTEL SCISSORS25DGS/S185MM	Yes

Prof. Dr. Muhammad Saleem
M.B.B.S., F.C.P.S., Gold Medalist
Head of Pediatric Department
Sheikh Zayed Medical College/Hospital
Rahim Yar Khan

DR. AKIF RAUF
Deputy Medical Superintendent
Sheikh Zayed Medical College/Hospital
Rahim Yar Khan

Dr. Zeeshan Iqbal
M.B.B.S., F.C.P.S.
Assistant Professor - anesthesia /ICU
Sheikh Zayed Medical College/Hospital
Rahim Yar Khan

ENG. AWAIS BIN TARIQ
Technical Officer
Sheikh Zayed Medical College/Hospital R.Y.K.

Head of Obs & Gynae Unit-II
S.Z.M.C/Hospital, Rahim Yar Khan

Sr.#	ITEM NO. 10E-Vascular Set (Local)	M/s Ashraf Nadeem
43	POTTS-DE MARTEL SCISSORS45DGS/S185MM	Yes
44	WEITLANER RETRACTOR 2X3T.SH.105MM	Yes
45	WEITLANER RETRACTOR 3X4T.SH.130MM	Yes
46	WEITLANER RETRACTOR 3X4T.SH.165MM	Yes
47	TC RYDER NEEDLE HOLDERDELSERR 135MM	Yes
48	TC RYDER NEEDLE HOLDERX-DELSERR175MM	Yes
49	TC RYDER NEEDLE HOLDERDELSERR 195MM	Yes
50	DUROGRIP NDL HLDR SERR 0.2/180MM	Yes
51	FERG-FRAZIER SUCT 6FR 2/110MM WRK-LGTH	Yes
52	DE'BAKEY ATR.BULLDOGCLAMP STR.30/90MM	Yes
53	DE'BAKEY ATR.BULLDOGCLAMP STR.45/105MM	Yes
54	DE'BAKEY ATR.BULLDOGCLAMP STR.65/125MM	Yes
55	PEAN-DE'BAKEY ATR.FRCPS STR155MM	Yes
56	DE'BAKEY-PEAN ATR.FRCPS STR.60MM 200MM	Yes
57	DE'BAKEY-PEAN ATR.FCPS STR89MM265MM	Yes
58	DE'BAKEY AORTA ANEURYSM CLMPCVD270MM	Yes
59	DE'BAKEY AORTA ANEURYSM CLAMPCVD330MM	Yes
60	DELICATE THUMB FORCEPS SMOOTH 200MM	Yes
61	STANDARD FORCEPS SERR 145MM	Yes
62	STANDARD FORCEPS SERR 200MM	Yes
63	METZENBAUM SCISSORS CVD 145MM	Yes
64	METZENBAUM SCISSORS CVD 200MM	Yes
65	UNIVERSAL SCISSORS BLACK 145MM	Yes
66	UNIVERSAL SCISSORS BLACK 180MM	Yes
67	YASARGIL/LEYLA SELF-RETAINING RETRI-ARM	Yes
68	BRAIN SPATULA FLAT MALL 11&13MM178MM	Yes
69	BRAIN SPATULA FLAT MALL15&18MM178MM	Yes
70	OLIVECRONA SPATULAFLATFLEX7&9MM178MM	Yes
71	OLIVECRONA BRAIN SPATULA 18 A.22MM	Yes
72	MAYO-HEGAR NEEDLE HOLDER150MM	Yes
73	MAYO-HEGAR NEEDLE HOLDER185MM	Yes
74	MAYO-HEGAR NEEDLE HOLDER200MM	Yes
75	HEGAR-MAYO NEEDLE HOLDER 265MM	Yes
76	TISSUE FORCEPS SERR CVD 140MM	Yes
77	TISSUE FORCEPS CVD 1X2 145MM	Yes
	Remarks	Responsive

Note:

Subject to approval of Sample.

Zeeshan
Dr. Zeeshan Khan
 M.B.S.S., F.C.P.S.
 Assistant Professor & Head of ICU
 Sh. Zayed Medical College Hospital
 Rahim Yar Khan

Eng. Anwar Bin Arif
Eng. ANWAR BIN ARIF
 Bio Medical Engineer/Technical Officer
 Sh. Zayed Medical College Hospital R.Y.K.

Prof. Dr. Anwar Rashed
Prof. Dr. Anwar Rashed
 Head of Obs & Gynae Unit-II
 S.Z.M.C Hospital, Rahim Yar Khan

Prof. Dr. Muhammad Saleem
Prof. Dr. Muhammad Saleem
 M.B.S.S., F.C.P.S. (Gen. Med.)
 Head of Pediatric Department
 Sh. Zayed Medical College/Hospital
 Rahim Yar Khan

Dr. Akif Rauf
DR. AKIF RAUF
 Deputy Medical Superintendent
 Sh. Zayed Medical College Hospital
 Rahim Yar Khan

Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan

Technical Evaluation Statement

ADP Scheme "Programme For Revamping of Emergencies of Tertiary Care Hospital In Punjab (Sheikh Zayed Hospital, Rahim Yar Khan)"

Sr.#	ITEM NO. 11-E Air Curtain (Local)	M/s Ashraf Nadeem
	KNOCK DOWN CRITERIA (COMMERCIAL EVALUATION)	
1		
2	Original Receipt for Purchase of Tender	Yes
3	An Affidavit (as per Specimen Performa on Stamp Paper of Rs. 100/-)	Yes
4	Copy of CDR/Bid Security	Yes
5	Bid Validity	Yes
6	Delivery Period	Yes
7	Valid Manufacturing Licence (Embassy Attested)	Local
	Remarks (Responsive/Non-Responsive)	Responsive
	Evulation parameter (quoted itmes/Equipment)	
1	Exclusive /Sole authorization agreement form the Foreign Manufacturer.	Local
2	Certificate from the Manufacturer that the after sales Services/backup Services shall be provided jointly with the local sole agent and incase of change of local agent, they will provide the after sale services themselfe for through newly oppointed agent for the period mentioned from the date of commissioning.	Local
3	Submission of valid legally enforceable exclusive authorization letter of manufacturer assuring full guarantee and warranty obligations as per enclosed manufacturer authorized form with the bid document.	Local
4	Valid Manufacturing License of quoted item Manufacturing embassy attested with certified documentary evidence of manufacturing site & location.	Local
5	The medical equipment offered from foreign countries of USA, Europe and Japan shall be eligible to participate and must bear FDA510k, CE(MDD) or MHLW (Ministry of Health, Labor and Welfare) standard.	Local
6	A Certificate from the manufacturer that the installation will be conducted in conformity with the system requirements by following the professional approach.	Local
7	Satisfactory Past Performance of the bidder for quoted producat	Local
8	The firms shall also declare the make, model, country of origin and country of manufacturing of all standard accessories to be provided with the equipment.	Local
9	The quoted model of imported product shall be available on the current official website of the manufacturer; otherwise the quoted product shall be considered obsolete/ redundant and will straight away be rejected. (Web link and Screenshot must be attached)	Local
10	The bid must comply with the required technical specifications of the quoted single item/ complete package. Incomplete offer will straightaway be rejected.	Local
11	Acceptance of terms and condition of tender documents duly signed and stamped	Local
	Remarks (Responsive/Non-Responsive)	Responsive
	Technical Evaluation Criteria Score	71

Dr. Zeesha Khan
M.B.B.S., F.C.P.S.
Assistant Professor & anesthiat ICU
Sheikh Zayed Medical College/Hospital
Rahim Yar Khan

Prof. Dr. Nuzhat Rasheed
Head of Obs & Gynae Unit-II
Sheikh Zayed Medical College/Hospital
Rahim Yar Khan

Prof. Dr. Muhammad Saleem
M.B.B.S., F.C.P.S. (Gen Med) (Gen Med)
Head of Pediatric Department
Sheikh Zayed Medical College/Hospital
Rahim Yar Khan

DR. AKIF RAUF
Deputy Medical Superintendent
Sheikh Zayed Medical College/Hospital
Rahim Yar Khan

Sr.#	ITEM NO. 11-E Air Curtain (Local)	M/s Ashraf Nadeem
	Technical Specification	
1	Electric Air Curtains to create an invisible downward airflow barrier to keep outside air separate from temperature controlled indoor air.	Yes
2	Reduce energy wastage and keep out pests, odours and dust at the same time	Yes
3	Rated Voltage Single Phase 220V~240V	Yes
4	Rated Frequency 50/60 Hz (60Hz is only for Single Phase 220V)	Yes
5	Supply Voltage Rated Voltage $\pm 10\%$	Yes
6	Operating Conditions Temperature : -10~40 °C Relative humidity : below 85%	Yes
7	RH Storage Conditions Temperature : -10~60 °C Relative humidity : 30~85% RH Motor Form	Yes
8	Form 4-pole motor, double shafts	Yes
9	Fan Category Cross-flow Impeller	Yes
	Remarks (Responsive/Non-Responsive)	Responsive

Note:

Subject to approval of Sample.

Zeesha
Dr. Zeesha Khan
 M.B.B.S., F.C.P.S.
 Assistant Professor Anaesthesia ICU
 Sheikh Zayed Medical College/Hospital
 Rahim Yar Khan

[Signature]

[Signature]
Dr. Zayed Hussain

[Signature]
DR. AKIF RAUF
 Deputy Medical Superintendent
 Sheikh Zayed Hospital Rahim Yar Khan

[Signature]
Prof. Dr. Nazhat Rashid
 Head of Obs & Gynae Unit-II
 S.Z.M.C Hospital, Rahim Yar Khan

[Signature]
Eng. AWAIS BIN TARIQ
 Bio Medical Engineer/Technical Officer
 Sh. Zayed Medical College/Hospital R.Y.K

[Signature]
Prof. Dr. Muhammad Saleem
 M.B.B.S., F.C.P.S., (Gold Medalist)
 Head of Pediatric Department
 S.Z. Zayed Medical College/Hospital
 Rahim Yar Khan

[Signature]

[Signature]
 Dr. Ashraf Nadeem
 M.B.B.S., F.C.P.S., (Gold Medalist)
 Head of Anaesthesia Unit
 Sheikh Zayed Medical College/Hospital
 Rahim Yar Khan

Sheikh Zayed Medical College / Hospital, Rahim Yar Khan
Technical Evaluation Statement (Financial Year 2021-22) on 28-10-2021

ADP Scheme "Programme for Revamping of Emergencies of Tertiary Care Hospital in Punjab (Sheikh Zayed Hospital, Rahim Yar Khan)

Sr.#	ITEM NO. 1L- Electronic Autoclave 1STU	M/S Sigma International	M/S Clinical Life	M/s Mediland Pakistan	M/s Total Technologies	M/s Total Technologies
	KNOCK DOWN CRITERIA (COMMERCIAL EVALUATION)					
1	Original Receipt for Purchase of Tender	Yes	Yes	Yes	Yes	Yes
2	An Affidavit (as per Specimen Performa on Stamp Paper of Rs. 100/-)	Yes	Yes	Yes	Yes	Yes
3	Copy of CDR/Bid Security	Yes	Yes	Bank Guarantee	Yes	Yes
4	Bid Validity	Yes	Yes	Yes	Yes	Yes
5	Delivery Period	Yes	Yes	Yes	Yes	Yes
6	Valid Manufacturing Licence (Embassy Attested)	Yes (Online Verified Registration Certificate)	Yes (Online Verified Registration Certificate)	Yes (Online Verified Registration Certificate)	Yes (Online Verified Registration Certificate)	Model not mentioned in Provided Ministry Letter & also not provided online verification of Registration Link
	Remarks (Responsive/Non-Responsive)	Responsive	Responsive	Responsive	Responsive	Non Responsive
	Evulation parameter (quoted itmes/Equipment)					
1	Exclusive /Sole authorization agreement form the Foreign Manufacturer.	Yes	Yes	Yes	Yes	Yes
2	Certificate from the Manufacturer that the after sales Services/backup Services shall be provided jointly with the local sole agent and incase of change of local agent, they will provide the after sale services themselves for through newly appointed agent for the period mentioned from the date of commissioning.	Yes	Yes	Yes	Yes	Yes
3	Submission of valid legally enforceable exclusive authorization letter of manufacturer assuring full guarantee and warranty obligations as per enclosed manufacturer authorized form with the bid document.	Yes	Yes	Yes	Yes	Yes
4	Valid Manufacturing License of quoted item Manufacturing embassy attested with certified documentary evidence of manufacturing site & location.	Yes	Yes	Yes	Yes	Model not mentioned in Provided Ministry Letter & also not provided online verification of Registration Link
5	The medical equipment offered from foreign countries of USA, Europe and Japan shall be eligible to participate and must bear FDA510k, CE(MDD) or MHLW (Ministry of Health, Labor and Welfare) standard.	Yes	Yes	Yes	Yes	Yes

Dr. Zeeshan Khan
 M.B.B.S., F.C.P.S.
 Assistant Professor Anaesthesia/ICU
 Sheikh Zayed Medical College/Hospital
 Rahim Yar Khan

Dr. AWAIS BIN TARIQ
 Bio Medical Engineer/Technical Officer
 Sh. Zayed Medical College/Hospital R.Y.K.

Dr. AWAIS BIN TARIQ
 Superintendent
 Rahim Yar Khan

Sr.#	ITEM NO. 1L- Electronic Autoclave 1STU	M/S Sigma International	M/S Clinical Life	M/s Mediland Pakistan	M/s Total Technologies	M/s Vertex Medical
6	A Certificate from the manufacturer that the installation will be conducted in conformity with the system requirements by following the professional approach.	Yes	Yes	Yes	Yes	Yes
7	Satisfactory Past Performance of the bidder for quoted product. (02 years)	Yes	Yes	Yes	Yes	Yes
8	The firms shall also declare the make, model, country of origin and country of manufacturing of all standard accessories to be provided with the equipment.	Yes	Yes	Yes	Yes	Yes
9	The quoted model of imported product shall be available on the current official website of the manufacturer; otherwise the quoted product shall be considered obsolete/ redundant and will straight away be rejected. (Web link and Screenshot must be attached)	Yes	Yes	Yes	Yes	Yes
10	The bid must comply with the required technical specifications of the quoted single item/ complete package. Incomplete offer will straightaway be rejected.	Yes	Yes	Yes	Yes	Yes
11	Acceptance of terms and condition of tender documents duly signed and stamped	Yes	Yes	Yes	Yes	Yes
	Remarks (Responsive/Non-Responsive)	Responsive	Responsive	Responsive	Responsive	Non Responsive
	Evaluation Criteria Score	75	77	83	84	81
	TECHNICAL SPECIFICATIONS					
	Make	Ceilitron	A.J.COSTA	Getinge Infection control	MMM GROUP	STEELCO, SPA
	Model	AC - 450	Steri 21 AIC 40466-1D GI	H533	UNISTERI SL336-1	VS 4/2 E/V
	Country of Manufacturer (Origin): USA, EU, Japan	Hungry	Portugal	Poland	Germany	ITALY
1	Fully automatic, programmable, microprocessor type	Yes	Yes	Yes	Yes	Yes
2	Touch screen colored display and integrated printer	Yes	Yes	Yes	Yes	Yes
3	Automatic motorized/pneumatic single door system	No	Yes	Yes	No	No
4	Time cycled	Yes	Yes	Yes	Yes	Yes
5	working pressure 32 psi	Yes	Yes	Yes	Yes	Yes
6	Safety interlock	Yes	Yes	Yes	Yes	Yes
7	Temperature & Pressure recorder	Yes	Yes	Yes	Yes	Yes
8	Chamber pressure indicator	Yes	Yes	Yes	Yes	Yes
9	Cycle indicator to determine the phase of sterilization cycle	Yes	Yes	Yes	Yes	Yes
10	Program/Cycle selection	Yes	Yes	Yes	Yes	Yes
11	Complete with standard accessories and removable shelves	Yes	Yes	Yes	Yes	Yes
12	capable of taking both packets and containers of all standard sizes	Yes	Yes	Yes	Yes	Yes
13	Chamber capacity 01 STU	Yes	Yes	Yes	Yes	Yes
14	Rectangular shape	Yes	Yes	Yes	Yes	Yes
15	Chamber, jacket and doors made of AISI 316 L/Ti	Yes	Yes	Yes	Yes	Yes
16	Complete system with built-in water saving system	Yes	Yes	Yes	Yes	Yes
17	Automatic heat exchanger and Air detector	Yes	Yes	Yes	Yes	Yes
18	Two loading / unloading trolleys and one loading carts compatible with system	Yes	Yes	Yes	Yes	Yes

Head of Gynae Unit
C/Hospital, Rahim Yar Khan

Asst. Prof. Dr. Zeeshan Khan
M.B.B.S., F.C.P.S.
Surgical & Obstetrician
Rahim Yar Khan

DR. AKIF RAUF
Deputy Medical Superintendent
Sheikh Zayed Hospital Rahim Yar Khan

ENG. ANAIS DIN TARIQ
Bio Medical Engineer
Sh. Zayed Medical Equipment Hospital

Prof. Dr. Muhammad Saleem
M.B.B.S., M.C.S., F.C.S., F.R.C.S. (Edinburgh)
Head of Pediatric Department
Sheikh Zayed Medical Equipment Hospital
Rahim Yar Khan

Sr.#	ITEM NO. 1L- Electronic Autoclave 1STU	M/S Sigma International	M/S Clinical Life	M/s Mediland Pakistan	M/s Total Technologies	M/s Vertex Medical
19	UPS of suitable capacity with minimum 15 minutes for Controller cum display for monitoring and controlling of parameters during power shedding provided/ installed by the manufacturer	Yes	Yes	Yes	Yes	Yes
20	RO Plant (50 gallon per day) as a part of standard specification (Local Made)	Yes	Yes	Yes	Yes	Yes
21	The firm should quote the price of RO plant with the sterilizer.	Yes	Yes	Yes	Yes	Yes
	Remarks	Non Responsive	Responsive	Responsive	Non Responsive	Responsive

Note:

- 1 M/S VERTEX MEDICAL is Non Responsive as per KNOCK DOWN CRITERIA.
- 2 M/S TOTAL TECHNOLOGIES is Non Responsive as per TECHNICAL CRITERIA.
- 3 M/S SIGMA INTERNATIONAL is Non Responsive as per TECHNICAL CRITERIA.

Zeesha
Dr. Zeeshan Khan
 M.B.B.S., F.C.P.S.
 Assistant Professor Anaesthesia ICU
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan

[Signature]
 S.Z.M. CH Hospital, Rahim Yar Khan.

[Signature]
 Deputy Medical Superintendent
 Sh. Zayed Medical College Hospital R.Y.K.

[Signature]
DR. AKIF RANEY
 Deputy Medical Superintendent
 Sheikh Zayed Hospital Rahim Yar Khan

[Signature]
Eng. AWAIS BIN TARIQ
 Bio Medical Engineer/Technical Officer
 Sh. Zayed Medical College Hospital R.Y.K.

[Signature]
 M. S. Khan
 M.B.B.S., F.C.P.S.
 Assistant Professor Anaesthesia ICU
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan

[Signature]
Prof. Dr. Muhammad Saleem
 M.B.B.S., F.C.P.S.
 Head of Department
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan

Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan
Technical Evaluation Statement (Financial Year 2021-22) on 28-10-2021

ADP Scheme "Programme For Revamping of Emergencies of Tertiary Care Hospital in Punjab (Sheikh Zayed Hospital, Rahim Yar Khan)"

Sr.#	ITEM NO. 2L-Mobile Operation Light	M/s Eastern Medical	M/s Radiant Medical	M/s KASBN INT.
	KNOCK DOWN CRITERIA (COMMERCIAL EVALUATION)			
1	Original Receipt for Purchase of Tender	Yes	Yes	Yes
2	An Affidavit (as per Specimen Performa on Stamp Paper of Rs. 100/-)	Yes	Yes	Yes
3	Copy of CDR/Bid Security	Yes	Yes	Yes
4	Bid Validity	Yes	Yes	Yes
5	Delivery Period	Yes	Yes	Yes
6	Valid Manufacturing Licence (Embassy Attested)	Not Provided any registration Certificate	Yes (Online Verified Registration Certificate)	Not provided English Version registration certificate & Online Link for Verification
	Remarks (Responsive/Non-Responsive)	Non Responsive	Responsive	Non Responsive
	Evaluation parameter (quoted item/Equipment)			
1	Exclusive /Sole authorization agreement from the Foreign Manufacturer.	Yes	Yes	Yes
2	Certificate from the Manufacturer that the after sales Services/backup Services shall be provided jointly with the local sole agent and in case of change of local agent, they will provide the after sale services themselves for through newly appointed agent for the period mentioned from the date of commissioning.	Yes	Yes	Yes
3	Submission of valid legally enforceable exclusive authorization letter of manufacturer assuring full guarantee and warranty obligations as per enclosed manufacturer authorized form with the bid document.	Yes	Yes	Yes
4	Valid Manufacturing License of quoted item Manufacturing embassy attested with certified documentary evidence of manufacturing site & location.	Not Provided any registration Certificate	Yes	Not provided English Version registration certificate & Online Link for Verification
5	The medical equipment offered from foreign countries of USA, Europe and Japan shall be eligible to participate and must bear FDA510k, CE(MDD) or MHLW (Ministry of Health, Labor and Welfare) standard.	Yes	Yes	Yes
6	A Certificate from the manufacturer that the installation will be conducted in conformity with the system requirements by following the professional approach.	Yes	Yes	Yes
7	Satisfactory Past Performance of the bidder for quoted product	Yes	Yes	Yes
8	The firms shall also declare the make, model, country of origin and country of manufacturing of all standard accessories to be provided with the equipment.	Yes	Yes	Yes

Dr. Zeb
 Assistant Professor Anesthesia ICU
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan

DR. AKIF RAUF
 Deputy Medical Superintendent
 Sheikh Zayed Hospital Rahim Yar Khan

ENG. WAIS BIN JARIG
 Bio Medical Engineer/ Technical Officer
 S. Zayed Medical College Hospital R.Y.K.

Prof. Dr. Muhammad Saleem
 M.D.S. MCPS, FCPS (Gold Medalist)
 Head of Pediatric Department
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan

Prof. Dr. Zahid Rashid
 Head of Obs & Gynae Unit-II
 S.Z.M.C Hospital, Rahim Yar Khan

Sr.#	ITEM NO. 2L-Mobile Operation Light	M/s Eastern Medical	M/s Radiant Medical	M/s KASBN INT.
9	The quoted model of imported product shall be available on the current official website of the manufacturer; otherwise the quoted product shall be considered obsolete/ redundant and will straight away be rejected. (Web link and Screenshot must be attached)	Yes	Yes	Yes
10	The bid must comply with the required technical specifications of the quoted single item/ complete package. Incomplete offer will straightaway be rejected.	Yes	Yes	Yes
11	Acceptance of terms and condition of tender documents duly signed and stamped	Yes	Yes	Yes
		Non Responsive	Responsive	Non Responsive
	Evaluation Criteria Score	85	79	76
	TECHNICAL SPECIFICATIONS			
	Make	Medifa	MEDICINOS	Ordisi
	Model	504220	LUMEN130	IG-100M
	Country of Manufacturer (Origin): USA, EU, Japan	Germany	LITHUANIA	Spain
1	Hermetically dust proof LED head.	Yes	Yes	Yes
2	Luminance at 1m distance 130,000 lux or above.	Yes	Yes	Yes
3	Color temperature 4000°-5000° degree kelvin or better.	Yes	Yes	Yes
4	Light field diameter between 15 – 30 cm or better (required lux should be available within max. light field area).	Yes	Yes	Yes
5	Color rendering index 94 or more.	Yes	Yes	Yes
6	LED life 50,000 hours or more.	Yes	Yes	Yes
7	Temp. rise in surgical surface <2° C	Yes	Yes	Yes
8	220 V / 50 Hz 1 phase.	Yes	Yes	Yes
9	Built-in rechargeable battery backup for at least 2 hours.	Yes	Yes	Yes
10	Autoclaveable handles.	Yes	Yes	Yes
	Remarks	Responsive	Responsive	Responsive

Note:

- 1 M/S EASTREN MEDICAL is Non Responsive as per KNOCK DOWN CRITERIA.
- 2 M/S KASBAN INTERNATIONAL is Non Responsive as per KNOCK DOWN CRITERIA.

Zeeshan
Dr. Zeeshan Khan
 M.B.B.S., F.C.P.S.
 Assistant Professor Anaesthesia/ICU
 Sheikh Zayed Medical College/Hospital
 Rahim Yar Khan

Prof. Dr. Nuzhat Rasheed
Prof. Dr. Nuzhat Rasheed
 Head of Obs & Gynae Unit-II
 S.Z.M.C Hospital, Rahim Yar Khan

Dr. Akif Rauf
DR. AKIF RAUF
 Deputy Medical Superintendent
 Sheikh Zayed Hospital Rahim Yar Khan

Eng. Awais Bin Tariq
Eng. AWAIS BIN TARIQ
 Bio Medical Engineer/Technical Officer
 Sh. Zayed Medical College/Hospital R.Y.K.

Prof. Dr. Muhammad Saleem
Prof. Dr. Muhammad Saleem
 MBBS, MCPS, FCPS (Gold Medalist)
 Head of Pediatric Department
 Sheikh Zayed Medical College/Hospital
 Rahim Yar Khan

Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan
Technical Evaluation Statement (Financial Year 2021-22) on 28-10-2021

ADP Scheme "Programme For Revamping of Emergencies of Tertiary Care Hospital In Punjab (Sheikh Zayed Hospital, Rahim Yar Khan)"

Sr.#	ITEM NO. 3-L OUTLET Forceps	M/s Med Impex
	KNOCK DOWN CRITERIA (COMMERCIAL EVALUATION)	
1	Original Receipt for Purchase of Tender	Yes
2	An Affidavit (as per Specimen Performa on Stamp Paper of Rs. 100/-	Yes
3	Copy of CDR/Bid Security	Yes
4	Bid Validity	Yes
5	Delivery Period	Yes
6	Valid Manufacturing Licence (Embassy Attested)	Local
	Remarks (Responsive/Non-Responsive)	Responsive
	Evaluation Criteria Score	74
	Evaluation parameter (quoted items/Equipment)	
1	Exclusive /Sole authorization agreement form the Foreign Manufacturer.	Local
2	Certificate from the Manufacturer that the after sales Services/backup Services shall be provided jointly with the local sole agent and in case of change of local agent, they will provide the after sale services themselves for through newly appointed agent for the period mentioned from the date of commissioning.	Local
3	Submission of valid legally enforceable exclusive authorization letter of manufacturer assuring full guarantee and warranty obligations as per enclosed manufacturer authorized form with the bid document.	Local
4	Valid Manufacturing License of quoted item Manufacturing embassy attested with certified documentary evidence of manufacturing site & location	Local
5	The medical equipment offered from foreign countries of USA, Europe and Japan shall be eligible to participate and must bear FDA510k, CE(MDD) or MHLW (Ministry of Health, Labor and Welfare) standard.	Local
6	A Certificate from the manufacturer that the installation will be conducted in conformity with the system requirements by following the professional approach.	Local
7	Satisfactory Past Performance of the bidder for quoted product	Local
8	The firms shall also declare the make, model, country of origin and country of manufacturing of all standard accessories to be provided with the equipment.	Local
9	The quoted model of imported product shall be available on the current official website of the manufacturer; otherwise the quoted product shall be considered obsolete/redundant and will straight away be rejected. (Web link and Screenshot must be attached)	Local
10	The bid must comply with the required technical specifications of the quoted single item/ complete package. Incomplete offer will straightaway be rejected.	Local
11	Acceptance of terms and condition of tender documents duly signed and stamped	Local
	Technical Specifications	
	wriggly outlet forceps	
	Remarks (Responsive/Non-Responsive)	Responsive

Note: Subject to approval of Sample

Dr. Zeesha Khan
 Assistant Professor
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan

Dr. Zeesha Khan
 Head of OHS & Gynaec Unit
 S.Z.M.C Hospital, Rahim Yar Khan

DR. AKIF RAUF
 Deputy Medical Superintendent
 Sheikh Zayed Hospital, Rahim Yar Khan

Eng. AWAIS BIN TARIQ
 Bio Medical Engineer/Technical Officer
 Sheikh Zayed Medical College Hospital

Prof. Dr. Muhammad Saleem
 Head of Pathology Department
 Sheikh Zayed Medical College Hospital

Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan
Technical Evaluation Statement (Financial Year 2021-22) on 28-10-2021

ADP Scheme "Programme For Revamping of Emergencies of Tertiary Care Hospital In Punjab (Sheikh Zayed Hospital, Rahim Yar Khan)"

Sr.#	ITEM NO. 4-L-Sonic Aid	M/s Med Impex	M/s Eastern Medical	M/s Vital Care
	KNOCK DOWN CRITERIA (COMMERCIAL EVALUATION)			
1	Original Receipt for Purchase of Tender	Yes	Yes	Yes
2	An Affidavit (as per Specimen Performa on Stamp Paper of Rs. 100/-	Yes	Yes	Yes
3	Copy of CDR/Bid Security	Yes	Yes	Yes
4	Bid Validity	Yes	Yes	Yes
5	Delivery Period	Yes	Yes	Yes
6	Valid Manufacturing Licence (Embassy Attested)	Yes (Online Verified Registration Certificate)	No Registration Certificate provided	Company profile page provided except registration certificate
	Remarks (Responsive/Non-Responsive)	Responsive	Non Responsive	Non Responsive
	Evaluation Criteria Score	74	85	77
	Evulation parameter (quoted Itmes/Equipmnet)			
1	Exclusive /Sole authorization agreement form the Foreign Manufacturer.	Yes	Yes	Yes
2	Certificate from the Manufacturer that the after sales Services/backup Services shall be provided jointly with the local sole agent and incase of change of local agent, they will provide the after sale services themselves for through newly appointed agent for the period mentioned from the date of commissioning.	Yes	Yes	Yes
3	Submission of valid legally enforceable exclusive authorization letter of manufacturer assuring full guarantee and warranty obligations as per enclosed manufacturer authorized form with the bid document.	Yes	Yes	Yes
4	Valid Manufacturing License of quoted item Manufacturing embassy attested with certified documentary evidence of manufacturing site & location.	Yes	No Registration Certificate provided	Company profile page provided except registration certificate
5	The medical equipment offered from foreign countries of USA, Europe and Japan shall be eligible to participate and must bear FDA510k, CE(MDD) or MHLW (Ministry of Health, Labor and Welfare) standard.	Yes	Yes	Yes
6	A Certificate from the manufacturer that the installation will be conducted in conformity with the system requirements by following the professional approach.	Yes	Yes	Yes
7	Satisfactory Past Performance of the bidder for quoted product	Yes	Yes	Yes
8	The firms shall also declare the make, model, country of origin and country of manufacturing of all standard accessories to be provided with the equipment.	Yes	Yes	Yes

Dr. Zeeshan Khan
 M.B.B.S., F.C.P.S.
 Assistant Professor Anaesthesia ICU
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan

Prof. Dr. Anwar Rashid
 Head of Ohs & Gynae Unit-II
 M.C.Hospital, Rahim Yar Khan

DR. AKTIF RAUF
 Deputy Medical Superintendent
 Sheikh Zayed Hospital Rahim Yar Khan

Engl. ANWAR BIN TARIQ
 Bio Medical Engineer/Technical Officer
 Sh. Zayed Medical College/Hospital R Y K

Prof. Dr. Muhammad Saleem
 M.B.B.S., F.C.P.S. (Gold Medalist)
 Head of Pediatric Department
 Sheikh Zayed Medical College/Hospital
 Rahim Yar Khan

Sr.#	ITEM NO. 4-L-Sonic Aid	M/s Med Impex	M/s Eastern Medical	M/s Vital Care
9	The quoted model of imported product shall be available on the current official website of the manufacturer; otherwise the quoted product shall be considered obsolete/ redundant and will straight away be rejected. (Web link and Screenshot must be attached)	Yes	Yes	Yes
10	The bid must comply with the required technical specifications of the quoted single item/ complete package. Incomplete offer will straightaway be rejected.	Yes	Yes	Yes
11	Acceptance of terms and condition of tender documents duly signed and stamped	Yes	Yes	Yes
	Remarks (Responsive/Non-Responsive)	Responsive	Non Responsive	Non Responsive
	TECHNICAL SPECIFICATIONS			
	Model	DD-901	DD-250	ES-102EX
	Make	Newman Medical	Ultrasound Technologies	Hadeco Inc
	Country of Manufacturer (Origin): USA, EU, Japan	USA	UK	Japan
	DESKTOP FETAL HEART RATE DETECTOR	-	-	-
1	Desktop type	Yes	Yes	Yes
2	LCD display with visualization of at least fetal heart rate.	Yes	Yes	Yes
3	Ultrasound working frequency in the range 2MHz or more	Yes	Yes	Yes
4	High sensitivity equipment compatible probe: 2 MHz or more	Yes	Yes	Yes
5	Heart rate measurement range not smaller than 50-210 bpm	Yes	Yes	Yes
6	Audio output reproduction of the fetal heart rate with integrated speaker and headphones.	Yes	Yes	Yes
7	Integrated Audio volume control system	Yes	Yes	Yes
8	Battery backup for 2hrs or more.	Yes	Yes	Yes
	Accessories:	Yes	Yes	Yes
	• 10 bottles of gel for patient application.	Yes	Yes	Yes
	• Local trolley with lockable casters	Yes	Yes	Yes
	Remarks	Responsive	Non Responsive	Non Responsive

Note:

- 1 M/S EASTREN MEDICAL is Non Responsive as per KNOCK DOWN CRITERIA.
- 2 M/S VITAL CARE is Non Responsive as per KNOCK DOWN CRITERIA.

Zeshan
Dr. Zeshan Khan
 MBBS, FRCGS, FRCR (Gold Medalist)
 Assistant Professor Amasheh Hospital
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan

Prof. Dr. Nuzhat Rashid
Prof. Dr. Nuzhat Rashid
 Head of GHS & Gynaecology Unit-II
 S.Z.M.C Hospital, Rahim Yar Khan

Dr. Akif Rauf
DR. AKIF RAUF
 Deputy Medical Superintendent
 Sheikh Zayed Hospital Rahim Yar Khan

Eng. Awais Bin Tariq
ENG. AWAIS BIN TARIQ
 Bio Medical Engineer/Technical Officer
 Sh. Zayed Medical College Hospital R.Y.K

Prof. Dr. Muhammad Saleem
Prof. Dr. Muhammad Saleem
 MBBS, MRCGS, FRCR (Gold Medalist)
 Head of Pediatric Department
 Sheikh Zayed Medical College Hospital
 Rahim Yar Khan