

Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan

Technical Evaluation Statement 27/05/22
 Purchase of Medical Equipments For Different Departments 2021-2022




ITEM NO. 1		M/S		M/S	
PATIENT MONITOR		RADIANT MEDICAL	SHIRAZI TRADING	IDEA MEDICAL	M/S
Knock Down Criteria Commercial Evaluation	Original Reciept of Tender	Yes	Yes	Yes	Yes
	Affidavit from Bidder	Yes	Yes	Yes	Yes
	Bid Security (CDR)	Yes	Yes	Yes	Yes
	Bid Validity (CDR)	Yes	Yes	Yes	Yes
	Delivery Period	Yes	Yes	Yes	Yes
Knock Down Criteria Vendor Evaluation	Sole Authorization/ Sole Agent Certificate	Yes	Yes	Not a sole distributor and RADIANT MEDICAL has also its sole distribution AGREEMENT and valid for three years.	Yes
	Technical & Engineering capability	Yes	Yes	Yes	Yes
	Certificate from the Manufacturer about the after sale service by Agent/ Itself	Yes	Yes	Yes	Yes
	Vendor Past Performance for Quoted Product	Yes	Yes	No past performance and orders of PaceTech Monitors	Yes
	Availability of relevant Tools and Testing /Calibration Equipment	Yes	Yes	Yes	Yes

Handwritten signature

Handwritten signature

Handwritten signature

PATIENT MONITOR	RADIANT MEDICAL	SHIRAZI TRADING	IDEA MEDICAL
Certificate FDA/ CE(MDD) Or MHLW	Yes	Yes	Latest FDA 510k certificate of PACETECH has no approval of Arrhythmia OR Alarm
Quoted Model Available on Website	Yes	Yes	Yes
Bid Comply with advertised Technical Specifications	Yes	Yes	Yes
Model	SC-1200	B125P	Vitalmax 4000
Make	Hunt Leigh Healthcare	GE-Healthcare	PaceTech
Country of Manufacturer: USA, EU, Japan	UK	USA	USA
Country of Manufacturing	UK	China	USA
Warranty (3 years)	Yes	Yes	Yes
Product Evaluation			
A	Operating Features and Characteristics:		
1	Bedside monitor for Adult and Paeds		
2	Non fade TFT, LCD color display	Yes	Yes
3	Electro-surgical interference suppression/protection	Yes	Yes
4	Defibrillator protection	Yes	Yes
5	Freeze and cascade facility.	Yes	Yes
6	Waveform trache spee: 25 & 50 mm/sec.	Yes	Yes
7	Screen size: min. 12" TFT/LCD Touch Screen color display.	Yes	Yes
B	Capacity to interface with LAN/WLAN for data transfer	Yes	Yes
	ECG :		




 2 of 4

PATIENT MONITOR		RADIANT MEDICAL	SHIRAZI TRADING	IDEA MEDICAL
1	Numeric: heart rate.	Yes	Yes	Yes
2	Waveform : Six waveform minimum, real time and freeze ECG trace	Yes	Yes	Yes
C NON-INVASIVE BLOOD PRESSURE (NIBP):				
1	Method: oscillometric principle	Yes	Yes	Yes
2	Numeric: systolic, diastolic and mean pressure	Yes	Yes	Yes
3	Selectable auto inflate interval settings	Yes	Yes	Yes
4	Rising cuff/continuous pressure display.	Yes	Yes	Yes
5	Reusable cuff	Yes	Yes	Yes
D TEMPERATURE:				
1	Numeric: temperature selectable in °C/°F.	Yes	Yes	Yes
E PULSE OXIMETRY:				
1	Numeric: 0-100% oxygen saturation measuring range.	Yes	Yes	Yes
2	Waveform-plethysmograph pulse.	Yes	Yes	Yes
3	Reusable sensor electrode.	Yes	Yes	Yes
F ARRHYTHMIA ANALYSIS:				
1	Arrhythmia analysis and ST analysis.	Yes	Yes	Yes
G RESPIRATION:				
1	Breath rate display and apnea alarms.	Yes	Yes	Yes
2	Sweep speed; 6.25, 12.5 mm/sec.	Yes	Yes	Yes
OTHER FEATURES:				
1	Trend data: Graphical and Tabular	Yes	Yes	Yes
ALARMS:				
1	High and Low (setable) on all parameters	Yes	Yes	Yes
2	Visual and Audible indication of alarm	Yes	Yes	Yes
OPERATING REQUIREMENTS:				
1	AC 220V/50 Hz	Yes	Yes	Yes
2	Built in rechargeable battery for atleast 1-2 hour AC power failure at full parameter	Yes	Yes	Yes
1	Built in rechargeable battery for atleast 1-2 hour AC power failure at full parameter	Yes	Yes	Yes
ACCESSORIES:				

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

PATIENT MONITOR		RADIANT MEDICAL	SHIRAZI TRADING	IDEA MEDICAL
1	Qty of Reusable sensors 3	Yes	Yes	Yes
2	ECG Leads 3 No.	Yes	Yes	Yes
3	NIBP Cuff 5 No	Yes	Yes	Yes
4	SPO2 Probe 3No	Yes	Yes	Yes
5	Wall mounted stand (Local)	Yes	Yes	Yes
BID STATUS		Responsive	Responsive	Non Responsive

Handwritten signature in blue ink.

Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan
Technical Evaluation Statement

ITEM NO. 2 ECG MACHINE		M/s HOSPICARE
knock Down Criteria Commercial Evaluation	Original Receipt of Tender	Yes
	Affidavit from Bidder	Yes
	Bid Security (CDR)	Yes
	Bid Validity	Yes
	Delivery Period	Yes
knock Down Criteria Vendor Evaluation	Sole Autorization/ Sole Agent Certificate	Yes
	Technical & Engineering capability	Yes
	Certificate from the Manufacturer about the after sale service by Agent/ Itself	Yes
	Certificate FDA/ CE(MDD) Or MHLW	Yes
	Vendor Past performance	Yes
	Availability of relevant Tools and Testing /Calibration Equipment	Yes
	Quoted Model Available on Website	Yes
	Bid Comply with advertised Technical Specifications	Yes
Product Evaluation	Model	ECG 100S
	Make	Cardioline
	Country of Manufacturer: USA, EU, Japan	Italy
	Country of Manufacturing	Italy
	Warranty (3 years)	Yes
ECG MACHINE (03 CHANNEL)		
1	Three channel ECG on at least 3-5 inches LCD display.	Yes
2	Display of six channel ECG simultaneously	Yes
3	Automatic Operation	Yes
4	Variable gain: 1/2, 1, 2 cm/mV	Yes
5	Thermal recorder for printing out six channels simultaneously.	Yes
6	Interpretation software.	Yes
7	Recording Trace speed: 10, 25 and 50 mm/sec,	Yes
8	Muscle artifact and AC (50Hz) interference filters	Yes
9	Defibrillator protection	Yes
10	Built-in AC operation & battery backup minimum 30mins	Yes
11	Paper size: 100-110mm	Yes

ECG MACHINE		HOSPICARE
12	Built-in AC interference, noise filter and baseline correction	Yes
13	Conductive Jell 1.0 L in small packing.	Yes
14	Paper Roll 20.	Yes
Accessories:		
1	Complete with standard accessories, including separate patient cables for Adult, Pediatric use with re-usable electrodes. No. of Electrodes 2 sets each. (Adult, Paeds)	Yes
2	Mobile Cart (Local) with two shelves having two lockable wheels at diagonal position.	Yes
BID STATUS		Responsive

The image contains four handwritten signatures in blue ink. The top-left signature is a large, stylized loop. The top-middle signature is a tall, vertical stroke with a loop at the top. The top-right signature is a large, horizontal loop with a smaller loop on top. The bottom signature is a more complex, multi-stroke signature with several loops and a horizontal line.

Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan
Technical Evaluation Statement

ITEM NO. 03		M/s
B.P APPARATUS (MERCURY TYPE)		RELIANCE TECHNOLOGY
Knock Down Criteria Commercial Evaluation	Original Receipt of Tender	Yes
	Affidavit from Bidder	Yes
	Bid Security (CDR)	Yes
	Bid Validity	Yes
	Delivery Period	Yes
	Vendor Past performance	Yes
	Availability of relevant Tools and Testing /Calibration Equipment	Yes
	Bid Comply with advertised Technical Specifications	Yes
Product Evaluation	Model	ALP K2
	Make	Japan
	Country of Manufacturer: USA, EU, Japan/Local	Japan
	Country of Manufacturing	Japan
	Warranty (03)	Yes
B.P APPARATUS MERCURY TYPE)		
1	Mercury 300mm Hg with die cast metal housing large reservoir with spilling over arrangement	Yes
2	tube with 3 mm silicone	Yes
3	Rubber bulbs. 01 Velcro cuff for Adults	Yes
4	Complete with basket for storage of cuffs	Yes
BID STATUS		Responsive

Subject to approval of SAMPLE by the Technical Committee

The image shows four distinct handwritten signatures in blue ink, arranged horizontally. The signatures are stylized and appear to be official approvals or endorsements.


Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan
Technical Evaluation Statement

ITEM NO.04 ULTRASONIC NEBULIZER		M/S MED IMPEX
knock Down Criteria Commercial Evaluation	Original Receipt of Tender	Yes
	Affidavit from Bidder	Yes
	Bid Security (CDR)	Yes
	Bid Validity	Yes
	Delivery Period	Yes
knock Down Criteria Vendor Evaluation	Sole Autorization/ Sole Agent Certificate	Yes
	Technical & Engineering capability	Yes
	Certificate from the Manufacturer about the after sale service by Agent/ Itself	Yes
	Certificate FDA/ CE(MDD) Or MHLW	Yes
	Vendor Past performance	Yes
	Availability of relevant Tools and Testing /Calibration Equipment	Yes
	Quoted Model Available on Website	Yes
	Bid Comply with advertised Technical Specifications	Yes
Product Evaluation	Model	Comfort 3000 ku-500
	Make	Koushin Industries
	Country of Manufacturer: USA, EU, Japan	Japan
	Country of Manufacturing	Japan
	Warranty (3 years)	Yes
ULTRASONIC NEBULIZER		
1	Ultrasonic energy for uniform and highly dense 1-5 microns	Yes
2	Medication cup with replaceable diaphragm	Yes
3	Made of highly resistant Sterilizable resin	Yes
4	Stand with solution bottle for safety	Yes
5	Nebulizing rate: 4 ml / min or greater	Yes
6	Mist particle size: Approx. 1-5 microns	Yes
7	Nebulizing time setting: 1-30 min or more adjustable	Yes
8	medication cup capacity: 150ml or better	
	Accessories :	Yes
1	Solution bottle 500 ml for approximate 4 hours continuous operation	Yes
2	Solution bottle support	Yes
3	Water Level controller	Yes
4	Trolley Pak Made	Yes
BID STATUS		Responsive

1 of 1

Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan
Technical Evaluation Statement

ITEM NO. 06		M/s
SUCTION MACHINE (LIGHT DUTY)		VITAL CARE
Knock Down Criteria Commercial Evaluation	Original Receipt of Tender	Yes
	Affidavit from Bidder	Yes
	Bid Security (CDR)	Yes
	Bid Validity	Yes
	Delivery Period	Yes
Knock Down Criteria Vendor Evaluation	Sole Authorization/ Sole Agent Certificate	Yes
	Technical & Engineering capability	Yes
	Certificate from the Manufacturer about the after sale service by Agent/ Itself	Yes
	Certificate FDA/ CE(MDD) Or MHLW	Yes
	Vendor Past performance	Yes
	Availability of relevant Tools and Testing /Calibration Equipment	Yes
	Quoted Model Available on Website	Yes
	Bid Comply with advertised Technical Specifications	Yes
Product Evaluation	Model	EUROVAC H-50
	Make	HERSILL
	Country of Manufacturer: USA, EU, Japan	Spain
	Country of Manufacturing	Spain
	Warranty (3 Year)	Yes
SUCTION MACHINE (LIGHT DUTY)		
1	Reciprocating, oil free pump mechanism.	Yes
2	Heavy duty Mobile Suction Unit with twin jars (Polysulfone / Polycarbonate type) of capacity up to 3 or 4 litter each, Autoclaveable.	Yes
3	Aspiration rate up to 20-30 litters/minutes or more at 650-900mm.Hg	Yes
4	Vacuum continuously adjustable	Yes
5	Foot vaccum regulator	
6	Triple flow safety device	Yes
7	Change over valve	Yes
8	Suction tubing of silicone with coupling connection for each jar	Yes
9	Noise Level 45 dB or less.	Yes
10	220V/50Hz.	Yes
Accessories:		
a	• 10 x bacterial filter	Yes
b	• Jars 2 extra	
c	• Original trolley with lockable wheels	Yes
BID STATUS		Responsive

Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan
Technical Evaluation Statement

ITEM NO. 07		M/s
EMERGENCY TROLLEY (LOCAL MADE)		RELIANCE TECHNOLOGY
Knock Down Criteria Commercial Evaluation	Original Receipt of Tender	Yes
	Affidavit from Bidder	Yes
	Bid Security (CDR)	Yes
	Bid Validity	Yes
	Delivery Period	Yes
	Vendor Past performance	Yes
	Bid Comply with advertised Technical Specifications	Yes
Product Evaluation	Model	Local
	Make	Local
	Country of Manufacturer: USA, EU, Japan/Local	Local
	Country of Manufacturing	Local
	Warranty (03 Years)	Yes
EMERGENCY TROLLEY (LOCAL MADE)		
1	Frame and carcass shall be constructed from mild steel powder coated/ Polymer Mounted on approx 125mm coaster.	Yes
2	The approximate internal dimensions of the medicine container shall be 80 w x 50 d x 30 h cm. (Approx.)	Yes
3	The hinged lid of the unit shall slope down to the front.	Yes
4	Lockable with heavy duty dead lock.	Yes
5	Shall be fitted with the following accessories: 4 tier rack for holding medications. Ampoule holding rack. Side mounted fold down writing flap. Different compartments for medicine with labeling. Lower shelf mounted under medicine container. Unit should be supplied with a security clamp for wall fixing at the appropriate location, thus ensuring that the trolley is secure when not in use.	Yes
BID STATUS		Responsive
Subject to approval of SAMPLE by the Technical Committee		

The image shows four distinct handwritten signatures in blue ink, arranged horizontally. The signatures are stylized and appear to be official approvals or endorsements related to the technical evaluation statement.

Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan
Technical Evaluation Statement

ITEM NO. 08		M/s
MEDICINE CABINET (LOCAL MADE)		RELIANCE TECHNOLOGY
knock Down Criteria Commercial Evaluation	Original Receipt of Tender	Yes
	Affidavit from Bidder	Yes
	Bid Security (CDR)	Yes
	Bid Validity	Yes
	Delivery Period	Yes
	Vendor Past performance	Yes
	Bid Comply with advertised Technical Specifications	Yes
Product Evaluation	Model	Local
	Make	Local
	Country of Manufacturer: Local/ USA, EU, Japan	Local
	Country of Manufacturing	Local
	Warranty (03 Years)	Yes
MEDICINE CABINET (LOCAL MADE)		
1	Guage 16	Yes
2	Metallic frame (CO ₂ Welding) with epoxy finish	Yes
3	The upper section has two metallic shelves&Two Glass doors (Lockable)	Yes
4	Glass thickness 8mm	Yes
5	The Lower section has one metallic shelf & two metallic doors (Lockable)	Yes
6	35 x 60 x 170 cm (Approx.)	Yes
7	Ground clearance about 5cm	Yes
BID STATUS		Responsive

Subject to Approval of SAMPLE by Technical Committee

Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan

Technical Evaluation Statement

ITEM NO. 09		M/s
(A) FOWLER BEDS (LOCAL MADE)		RELIANCE TECHNOLOGY
Knock Down Criteria Commercial Evaluation	Original Receipt of Tender	Yes
	Affidavit from Bidder	Yes
	Bid Security (CDR)	Yes
	Bid Validity	Yes
	Delivery Period	Yes
	Technical & Engineering capability	Yes
	Vendor Past performance	Yes
	Bid Comply with advertised Technical Specifications	Yes
Product Evaluation	Model	Local
	Make	Local
	Country of Manufacturer: USA, EU, Japan/Local	Local
	Country of Manufacturing	Local
	Warranty (03 Years)	Yes
FOWLER BEDS (SEMI FOWLER) LOCAL MADE		
1	Rectangular Frame 18 SWG	Yes
2	Laying Area Made of reinforced Steel Strip of 3" x 1" x 1/2 20SWG	Yes
3	Crank Screw made of 7/8 x 14" Steel rod with square type threading and locking mechanism for extreme position.	Yes
4	Laying plat from, top having Semi fowler adjustment.	Yes
5	Mobile on four good quality castors 125mm (two diagonal front wheel lockable fitted under legs head and foot panel. First quality with washable Rexene cover	Yes
BID STATUS		Responsive

Subject to approval of Sanmple by the Technical committee

Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan
Technical Evaluation Statement

ITEM NO. 09		M/s
(B) BEDSIDE LOCKER (LOCAL MADE)		RELIANCE TECHNOLOGY
Knock Down Criteria Commercial Evaluation	Original Receipt of Tender	Yes
	Affidavit from Bidder	Yes
	Bid Security (CDR)	Yes
	Bid Validity	Yes
	Delivery Period	Yes
	Vendor Past performance	Yes
	Bid Comply with advertised Technical Specifications	Yes
Product Evaluation	Make	Local
	Country of Manufacturer: USA, EU, Japan/Local	Local
	Country of Manufacturing	Local
	Warranty (03 Years)	Yes
	BEDSIDE LOCKER (LOCAL MADE)	
1	Powder coated sheet steel construction.	Yes
2	Mild steel sheet 20 SWG	Yes
3	Two drawers with one bottom compartment	Yes
4	Built-in handles for durability	Yes
5	Plastic top laminated with raised edges on 3 side's	Yes
6	75 mmm twin castors rust proof, water proof ,all lockables.	Yes
6	Castors with plastic tires for noise free operation on all floors	
7	All metal parts should be welded by CO2 gas	Yes
8	Approx. overall size: 18" W x 15" D x 34" H (Approx)	Yes
BID STATUS		Responsive

Subject to Approval of Sample by the Technical committee

The image shows four distinct handwritten signatures in blue ink, positioned below the technical evaluation statement. The signatures are stylized and appear to be from different individuals.

Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan
Technical Evaluation Statement

ITEM NO. 09 (C)IV STAND (LOCAL MADE)		M/s RELIANCE TECHNOLOGY
knock Down Criteria Commercial Evaluation	Original Receipt of Tender	Yes
	Affidavit from Bidder	Yes
	Bid Security (CDR)	Yes
	Bid Validity	Yes
	Delivery Period	Yes
	Technical & Engineering capability	Yes
	Vendor Past performance	Yes
	Bid Comply with advertised Technical Specifications	Yes
Product Evaluation	Model	Local
	Make	Local
	Country of Manufacturer: USA, EU, Japan/Local	Local
	Country of Manufacturing	Local
	Warranty(03 Years)	Yes
IV STAND (LOCAL MADE)		
1	Base made of Mild steel Strip 1.5 width, with 2" casters. Central rod of 1" round pipe Mild steel.	Yes
2	The height adjustable pipe made of ¾" tubular stainless steel with four prongs rod of SS 4mm for infusion bags	Yes
3	Heavy Base	Yes
BID STATUS		Responsive
Subject to approval of Sample by the Technical Committee		

The image shows four distinct handwritten signatures in blue ink, positioned below the table. The signatures are stylized and appear to be official approvals or endorsements related to the technical evaluation process.

Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan
Technical Evaluation Statement

ITEM NO. 09		M/s
(D) OVER BED TABLE (LOCAL MADE)		RELIANCE TECHNOLOGY
knock Down Criteria Commercial Evaluation	Original Receipt of Tender	Yes
	Affidavit from Bidder	Yes
	Bid Security (CDR)	Yes
	Bid Validity	Yes
	Delivery Period	Yes
	Vendor Past performance	Yes
	Bid Comply with advertised Technical Specifications	Yes
Product Evaluation	Model	Local
	Make	Local
	Country of Manufacturer: USA, EU, Japan/Local	Local
	Country of Manufacturing	Local
	Warranty (03 Years)	Yes
OVER BED TABLE (LOCAL MADE)		
1	Power coated of mild steel construction	Yes
2	stainless steel internal column	Yes
3	Gas spring assisted hieght	Yes
4	Height adjustable from 70 to 100cm adjustment	Yes
5	Plastic top laminated with rounded edges and cup/glass holders	Yes
6	Approx top size: 80*32 cm (Approx)	Yes
7	50mm twin casters	Yes
8	Casters with rubber tires for noise free operation on floors	Yes
BID STATUS		Responsive
Subject to approval of SAMPLE by the Technical Committee		






Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan
Technical Evaluation Statement

ITEM NO. 09		M/s
(E) ATTENDANT BENCH (LOCAL MADE)		RELIANCE TECHNOLOGY
Knock Down Criteria Commercial Evaluation	Original Receipt of Tender	Yes
	Affidavit from Bidder	Yes
	Bid Security (CDR)	Yes
	Bid Validity	Yes
	Delivery Period	Yes
	Vendor Past performance	Yes
	Bid Comply with advertised Technical Specifications	Yes
Product Evaluation	Model	Local
	Make	Local
	Country of Manufacturer: USA, EU, Japan/Local	Local
	Country of Manufacturing	Local
	Warranty (03 Years)	Yes
(E) ATTENDANT BENCH (LOCAL MADE)		
1	Ms pipe/tube base structure (1" x 1½") with aluminium paint.	Yes
2	Full Wooden top of rose wood ¾" thick (best quality)	Yes
3	Size of the bench is 60"*18"*15" (L*W*H)	Yes
BID STATUS		Responsive
Subject to approval of SAMPLE by the Technical Committee		

**Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan
Technical Evaluation Statement**

ITEM NO.10		M/s
OXYGEN FLOWMETER (WALL TYPE)		RELIANCE TECHNOLOGY
Knock Down Criteria Commercial Evaluation	Original Receipt of Tender	Yes
	Affidavit from Bidder	Yes
	Bid Security (CDR)	Yes
	Bid Validity	Yes
	Delivery Period	Yes
	Technical & Engineering capability	Yes
	Vendor Past performance	Yes
	Bid Comply with advertised Technical Specifications	Yes
Product Evaluation	Model	HP906F/BS 300ml
	Make	Imported
	Country of Manufacturer: USA, EU, Japan/Local	Imported
	Country of Manufacturing	China
	Warranty(03 Years)	Yes
OXYGEN FLOWMETER (WALLTYPE)		Yes
BID STATUS		Responsive
Subject to Approval of SAMPLE by the Technical Committee		

The image shows four distinct handwritten signatures in blue ink, arranged in two rows. The top row contains three signatures, and the bottom row contains one signature.

Sheikh Zayed Medical College/ Hospital, Rahim Yar Khan
 Technical Evaluation Statement

ITEM NO. 11 DEFIBRILLATOR	M/S		M/S
	MEDEQUIP & SYSTEMS	RADIANT Medical	
Knock Down Criteria Commercial Evaluation Vendor Evaluation Knock Down Criteria	Original Receipt of Tender	Yes	Yes
	Affidavit from Bidder	Yes	Yes
	Bid Security (CDR)	Yes	Yes
	Bid Validity	Yes	Yes
	Delivery Period	Yes	Yes
	Sole Authorization/ Sole Agent Certificate	Yes	Yes
	Technical & Engineering capability	Yes	Yes
	Certificate from the Manufacturer about the after sale service by Agent/ Itself	Yes	Yes
	Certificate FDA/ CE(MDD) Or MHLW	Yes	Yes
	Vendor Past performance	Yes	Yes
Product Evaluation	Availability of relevant Tools and Testing /Calibration Equipment	Yes	Yes
	Quoted Model Available on Website	Yes	Yes
	Bid Comply with advertised Technical Specifications	Yes	Yes
	Model	M2	REANIBEX-700
	Make	ZOLL	OSATU
	Country of Manufacturer: USA, EU, Japan	USA	Spain
	Country of Manufacturing	CHINA	Spain
	Warranty (03)	Yes	Yes
	Biphasic transthoracic (external) defibrillator with LCD colour display	Yes	Yes
	Synchronized output with ECG	Yes	Yes





 1 of 2

ITEM NO. 11 DEFIBRILLATOR		M/S	
		MEDEQUIP & SYSTEMS	RADIANT Medical
3	Energy selection & delivery on control panel and paddles for external defibrillation	Yes	Yes
4	Energy selection and delivery on control panel for internal defibrillation.	Yes	Yes
5	Charging Indicator	Yes	Yes
6	The energy range should be adjustable for peds and adults up to 200Joules	Yes	Yes
7	Charging Time for full energy should be less than 05 sec	Yes	Yes
8	Screen Size of approx. 5 inch colored	Yes	Yes
9	Display of HR, ECG through paddles and Lead I,II & III patient cable	Yes	Yes
10	Built in recorder for printing of full summary on standard 50mm paper	Yes	Yes
11	Alarms for High and low Heart rate, low battery warning	Yes	Yes
12	Built-in Rechargeable battery with charger for minimum 50 shocks at max energy	Yes	Yes
13	Auto tester/self-check	Yes	Yes
14	External Paddles (Adult, Paed, Neonate)	Yes	Yes
15	AED facility with cable	Yes	Yes
16	Pacing facility	Yes	Yes
17	AC 220V / 50Hz operated	Yes	Yes
	Accessories:	-	-
	Complete with standard accessories, including reusable type	Yes	Yes
1	Adult, Pediatric & Neonatal sensors	Yes	Yes
2	Original trolley/cart	Yes	Yes
3	Qty of Reusable sensors 2 No	Yes	Yes
4	Internal Paddle(Adult, Paed, Neonate)	Yes	Yes
7	Spo2 Qty 2 No.	Yes	Yes
8	Pacing pads Qty 2 Nos.	Yes	Yes
BID STATUS		RESPONSIVE	RESPONSIVE



 2 of 2